



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB100251

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP70701
COD calibration std. 100 ppm	WP70705
COD calibration std. 10 ppm	WP70703
COD calibration std. 150 ppm	WP70706
COD calibration std. 50 ppm	WP70704
COD calibration std. 0 ppm	WP70698
COD Digestion Vials Low Level 0-150Mg/L	W2442
COD ICV-LCS std, 50ppm	WP70807
COD CCV std, 50ppm	WP70808

Temp In (C): 150	Date In: 01/10/2019	Time In: 11:30
Temp Out (C): 150	Date Out: 01/10/2019	Time Out: 13:30

Intercept: -0.6499

Slope: 1.004

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	01/05/2019	14:00
2	CAL2	10	1	Water	9.000	01/05/2019	14:00
3	CAL3	50	1	Water	49.000	01/05/2019	14:01
4	CAL4	100	1	Water	100.000	01/05/2019	14:01
5	CAL5	150	1	Water	150.000	01/05/2019	14:02



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB100251

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	50.448	01/05/2019	14:02
2	ICB				1	Water	0.000	0.647	01/05/2019	14:03
3	CCV1	50			1	Water	51.000	51.444	01/10/2019	14:30
4	CCB1				1	Water	0.000	0.647	01/10/2019	14:30
5	LB100251BL				1	Water	0.000	0.647	01/10/2019	14:31
6	LB100251BS	50			1	Water	49.000	49.452	01/10/2019	14:31
7	K1085-01				1	Water	5.000	5.627	01/10/2019	14:32
8	K1085-01DUP				1	Water	6.000	6.623	01/10/2019	14:32
9	K1085-01MS	50			1	Water	54.000	54.432	01/10/2019	14:33
10	K1085-01MSD	50			1	Water	55.000	55.428	01/10/2019	14:33
11	K1086-01				1	Water	136.000	136.105	01/10/2019	14:34
12	K1100-09				1	Water	79.000	79.333	01/10/2019	14:34
13	CCV2	50			1	Water	51.000	51.444	01/10/2019	14:35
14	CCB2				1	Water	0.000	0.647	01/10/2019	14:35

WORKLIST(Hardcopy Internal Chain)

LB100 251

WorkList Name : COD011019

WorkList ID : 121200

Date : 1/10/2019 10:46:13 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/21/2019	Water	K1085-01	COD	Conc H2SO4 1c	NATI04	F12	DSN001	01/08/2019	SM5220 D
01/11/2019	Water	K1086-01	COD	Conc H2SO4 1c	PSEG01	E61	CF-1819-W	01/08/2019	SM5220 D
01/14/2019	Water	K1100-09	COD	Conc H2SO4 1c	PSEG01	E61	WATER-DRUMS-A	01/09/2019	SM5220 D

Date/Time 01/10/19 10:51
 Received by: IB
 Relinquished by: CQ

Date/Time 01/10/19 13:00
 Received by: CQ
 Relinquished by: IB