



PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 1/11/2019

OVENTEMP IN Celsius(°C): 108  
Time IN: 17:20  
In Date: 01/10/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:11  
Out Date: 01/11/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-1

QC:LB100253

| Lab ID   | Client Sample ID      | Dish# | Dish Wt(g)<br>(A) | Dish +<br>Sample Wt(g)<br>(B) | Dish + Dry<br>Sample Wt(g)<br>(C) | % Solid |
|----------|-----------------------|-------|-------------------|-------------------------------|-----------------------------------|---------|
| K1011-01 | OR-02-011019-A        | 1     | 1.14              | 9.84                          | 9.00                              | 90.3    |
| K1011-03 | OR-02-011019-B        | 2     | 1.13              | 9.86                          | 8.97                              | 89.8    |
| K1082-27 | SP-1-1                | 3     | 1.15              | 9.96                          | 7.7                               | 74.3    |
| K1082-28 | SP-1-2                | 4     | 1.19              | 9.98                          | 8.8                               | 86.6    |
| K1082-29 | SP-1-3                | 5     | 1.14              | 9.66                          | 8.36                              | 84.7    |
| K1082-30 | SP-1-4                | 6     | 1.14              | 9.61                          | 8.62                              | 88.3    |
| K1082-31 | SP-1-5                | 7     | 1.13              | 9.54                          | 9.00                              | 93.6    |
| K1082-32 | SP-1-6                | 8     | 1.18              | 9.93                          | 8.58                              | 84.6    |
| K1102-01 | CPSB-18 (15-20)       | 9     | 1.17              | 9.87                          | 7.86                              | 76.9    |
| K1102-02 | CPSB-18 (20-25)       | 10    | 1.16              | 9.77                          | 5.05                              | 45.2    |
| K1102-03 | CPSB-02 (15-20)       | 11    | 1.15              | 9.96                          | 9.07                              | 89.9    |
| K1102-04 | CPSB-02 (20-25)       | 12    | 1.19              | 9.8                           | 4.93                              | 43.4    |
| K1102-05 | CPSB-13 (10-15)       | 13    | 1.17              | 9.88                          | 7.82                              | 76.3    |
| K1102-06 | CPSB-13 (15-20)       | 14    | 1.11              | 9.52                          | 3.71                              | 30.9    |
| K1102-07 | CPSB-14 (10-15)       | 15    | 1.16              | 9.79                          | 6.47                              | 61.5    |
| K1102-08 | CPSB-14 (15-20)       | 16    | 1.18              | 9.98                          | 9.08                              | 89.8    |
| K1112-01 | DRUM                  | 17    | 1.00              | 2.00                          | 2.00                              | 100.0   |
| K1112-02 | TOTES-1-2             | 18    | 1.00              | 2.00                          | 2.00                              | 100.0   |
| K1113-01 | BCB-COMPOSITE         | 19    | 1.17              | 9.94                          | 9.7                               | 97.3    |
| K1113-02 | BCB-3-1500            | 20    | 1.12              | 9.56                          | 9.28                              | 96.7    |
| K1114-01 | PM-A                  | 21    | 1.00              | 2.00                          | 2.00                              | 100.0   |
| K1114-02 | PM-B                  | 22    | 1.00              | 2.00                          | 2.00                              | 100.0   |
| K1114-03 | PM-C                  | 23    | 1.00              | 2.00                          | 2.00                              | 100.0   |
| K1114-04 | PM-D                  | 24    | 1.00              | 2.00                          | 2.00                              | 100.0   |
| K1114-05 | PM-E                  | 25    | 1.00              | 2.00                          | 2.00                              | 100.0   |
| K1114-06 | PM-F                  | 26    | 1.00              | 2.00                          | 2.00                              | 100.0   |
| K1115-01 | S72-IDW01             | 27    | 1.16              | 9.93                          | 8.38                              | 82.3    |
| K1116-01 | S49-IDW03             | 28    | 1.19              | 9.78                          | 8.86                              | 89.3    |
| K1117-01 | S32-IDW22             | 29    | 1.19              | 9.8                           | 8.57                              | 85.7    |
| K1118-01 | FENCE-SOIL-PILE-1     | 30    | 1.15              | 9.95                          | 9.74                              | 97.6    |
| K1118-02 | FENCE-SOIL-PILE-1-1   | 31    | 1.15              | 9.95                          | 9.74                              | 97.6    |
| K1118-03 | FENCE-SOIL-PILE-1-22  | 32    | 1.15              | 9.95                          | 9.74                              | 97.6    |
| K1118-04 | FENCE-SOIL-PILE-1     | 33    | 1.15              | 9.95                          | 9.74                              | 97.6    |
| K1118-05 | TESTPIT-SOIL-PILE-1   | 34    | 1.14              | 9.61                          | 8.63                              | 88.4    |
| K1118-06 | TESTPIT-SOIL-PILE-1-1 | 35    | 1.14              | 9.61                          | 8.63                              | 88.4    |
| K1118-07 | TESTPIT-SOIL-PILE-1-2 | 36    | 1.14              | 9.61                          | 8.63                              | 88.4    |
| K1118-08 | TESTPIT-SOIL-PILE-1   | 37    | 1.14              | 9.61                          | 8.63                              | 88.4    |
| K1118-09 | CONCRETE-PILE         | 38    | 1.00              | 2.00                          | 2.00                              | 100.0   |
| K1119-01 | DS-A                  | 39    | 1.13              | 9.52                          | 8.09                              | 83.0    |



# PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 1/11/2019

OVENTEMP IN Celsius(°C): 108  
Time IN: 17:20  
In Date: 01/10/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:11  
Out Date: 01/11/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-1

QC:LB100253

| Lab ID   | Client Sample ID | Dish# | Dish Wt(g)<br>(A) | Dish +<br>Sample Wt(g)<br>(B) | Dish + Dry<br>Sample Wt(g)<br>(C) | % Solid |
|----------|------------------|-------|-------------------|-------------------------------|-----------------------------------|---------|
| K1119-03 | DS-B             | 40    | 1.16              | 9.67                          | 8.24                              | 83.2    |
| K1119-05 | DS-C             | 41    | 1.14              | 9.95                          | 8.46                              | 83.1    |
| K1119-07 | DS-D             | 42    | 1.18              | 9.87                          | 8.49                              | 84.1    |
| K1122-01 | SB-1I (17-17.5)  | 43    | 1.18              | 9.98                          | 9.28                              | 92.0    |
| K1122-02 | SB-2I (16-16.5)  | 44    | 1.17              | 9.69                          | 8.88                              | 90.5    |
| K1122-03 | SB-3I (4-4.5)    | 45    | 1.18              | 9.63                          | 8.51                              | 86.7    |
| K1122-04 | SB-3I (16-16.5)  | 46    | 1.11              | 9.81                          | 9.14                              | 92.3    |
| K1122-05 | SB-4I (4-4.5)    | 47    | 1.19              | 9.94                          | 9.22                              | 91.8    |
| K1122-07 | SB-5I (15.5-16)  | 48    | 1.13              | 9.61                          | 9.03                              | 93.2    |
| K1122-08 | SB-6I (9-9.5)    | 49    | 1.1               | 9.59                          | 8.94                              | 92.3    |
| K1122-10 | SB-7I (16-16.5)  | 50    | 1.13              | 9.6                           | 9.03                              | 93.3    |
| K1122-11 | SB-8I (16-16.5)  | 51    | 1.14              | 9.84                          | 8.95                              | 89.8    |
| K1124-01 | ER-SOIL          | 52    | 1.18              | 9.93                          | 8.15                              | 79.7    |
| K1124-02 | ER-SOIL          | 53    | 1.18              | 9.93                          | 8.15                              | 79.7    |
| K1124-03 | T-20             | 54    | 1.00              | 2.00                          | 2.00                              | 100.0   |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$