


LB100262

INSTRUMENT ID: CV1

7470A

μ Abs.:

40673



Linear

A= 0.0000e+000

B= 2.4517e-004

C= 5.8123e-003

Rho= 0.9997832

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.004	0.004	-9	0.000	-9				
0.2	0.200	0.171	-0.029	674	0.0 %	674				
2.5	2.500	2.450	-0.050	9971	0.0 %	9971				
5.0	5.000	5.165	0.165	21042	0.0 %	21042				
7.5	7.500	7.433	-0.067	30294	0.0 %	30294				
10.0	10.000	9.977	-0.023	40673	0.0 %	40673	0			

LB100262 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Jan 2019 10:29:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	-9	-	0.00007470A	PPB		11 Jan 2019 10:40:31	S	Std
0.2 - 1	50-2	674	-	0.20007470A	PPB		11 Jan 2019 10:42:39	S	Std
2.5 - 1	52-5	9971	-	2.50007470A	PPB		11 Jan 2019 10:44:48	S	Std
5.0 - 1	55-0	21042	-	5.00007470A	PPB		11 Jan 2019 10:46:56	S	Std
7.5 - 1	57-5	30294	-	7.50007470A	PPB		11 Jan 2019 10:49:04	S	Std
10.0 - 1	510-0	40673	-	10.00007470A	PPB		11 Jan 2019 10:51:12	S	Std
ICV57 - 1	ICV57	15723	3.8606	-7470A	PPB		11 Jan 2019 10:55:08	U	SMPL
ICB57 - 1	ICB57	-50	-0.0064	-7470A	PPB		11 Jan 2019 10:57:15	U	SMPL
CCV16 - 1	CCV16	21016	5.1582	-7470A	PPB		11 Jan 2019 10:59:22	U	SMPL
CCB16 - 1	CCB16	-57	-0.0082	-7470A	PPB		11 Jan 2019 11:01:28	U	SMPL
CRA - 1	CRA	763	0.1929	-7470A	PPB		11 Jan 2019 11:03:35	U	SMPL
HighStd - 1	HighStd	40286	9.8826	-7470A	PPB		11 Jan 2019 11:05:42	U	SMPL
ChkStd - 1	ChkStd	28984	7.1117	-7470A	PPB		11 Jan 2019 11:07:49	U	SMPL
PB116312BL - 1	PBW	-33	-0.0023	-7470A	PPB		11 Jan 2019 11:09:57	U	SMPL
PB116312BS - 1	LCSW	14522	3.5661	-7470A	PPB		11 Jan 2019 11:12:05	U	SMPL
K1005-02 - 1	HR-06-010918-A	-17	0.0016	-7470A	PPB		11 Jan 2019 11:14:13	U	SMPL
K1005-02DUP - 1	HR-06-010918-ADUP	36	0.0146	-7470A	PPB		11 Jan 2019 11:16:21	U	SMPL
K1005-02MS - 1	HR-06-010918-AMS	14839	3.6438	-7470A	PPB		11 Jan 2019 11:18:30	U	SMPL
K1005-02MSD - 1	HR-06-010918-AMSD	13814	3.3925	-7470A	PPB		11 Jan 2019 11:20:38	U	SMPL
K1005-04 - 1	HR-06-010918-B	146	0.0416	-7470A	PPB		11 Jan 2019 11:22:45	U	SMPL
CCV17 - 1	CCV17	20749	5.0928	-7470A	PPB		11 Jan 2019 11:24:51	U	SMPL
CCB17 - 1	CCB17	-66	-0.0104	-7470A	PPB		11 Jan 2019 11:26:59	U	SMPL
K1005-06 - 1	HR-06-010918-C	110	0.0328	-7470A	PPB		11 Jan 2019 11:29:06	U	SMPL
K1015-02 - 1	SU-03-010918-A	105	0.0316	-7470A	PPB		11 Jan 2019 11:31:13	U	SMPL
K1015-04 - 1	SU-03-010918-B	76	0.0244	-7470A	PPB		11 Jan 2019 11:33:20	U	SMPL
K1083-03 - 1	OILY-STONES	135	0.0389	-7470A	PPB		11 Jan 2019 11:35:27	U	SMPL
K1087-17 - 1	MH-444-C	133	0.0384	-7470A	PPB		11 Jan 2019 11:37:35	U	SMPL
K1093-01 - 1	METUCHEN-DRUM-1-1	91	0.0281	-7470A	PPB		11 Jan 2019 11:39:43	U	SMPL
K1094-02 - 1	CROSSWICKS-DRUM-1	140	0.0401	-7470A	PPB		11 Jan 2019 11:41:50	U	SMPL
K1094-04 - 1	CROSSWICKS-DRUM-2	51	0.0183	-7470A	PPB		11 Jan 2019 11:43:59	U	SMPL
K1099-04 - 1	TP-15	88	0.0274	-7470A	PPB		11 Jan 2019 11:46:08	U	SMPL
K1099-08 - 1	TP-16	78	0.0249	-7470A	PPB		11 Jan 2019 11:48:15	U	SMPL
CCV18 - 1	CCV18	21135	5.1874	-7470A	PPB		11 Jan 2019 11:50:22	U	SMPL
CCB18 - 1	CCB18	-67	-0.0106	-7470A	PPB		11 Jan 2019 11:52:29	U	SMPL
K1099-12 - 1	TP-17-14	8	0.0078	-7470A	PPB		11 Jan 2019 11:54:36	U	SMPL
K1099-16 - 1	TP-18	45	0.0168	-7470A	PPB		11 Jan 2019 11:56:43	U	SMPL
K1099-20 - 1	TP-19-20	0	0.0058	-7470A	PPB		11 Jan 2019 11:58:51	U	SMPL
K1099-24 - 1	TP-21	718	0.1818	-7470A	PPB		11 Jan 2019 12:00:58	U	SMPL
K1100-02 - 1	SOIL-DRUM-A	-9	0.0036	-7470A	PPB		11 Jan 2019 12:03:05	U	SMPL
K1100-04 - 1	SOIL-DRUM-B	14	0.0092	-7470A	PPB		11 Jan 2019 12:05:13	U	SMPL
K1100-06 - 1	SOIL-DRUM-C	0	0.0058	-7470A	PPB		11 Jan 2019 12:07:21	U	SMPL
K1100-08 - 1	SOIL-DRUM-D	17	0.0100	-7470A	PPB		11 Jan 2019 12:09:30	U	SMPL
PB116313BL - 1	PBW	22	0.0112	-7470A	PPB		11 Jan 2019 12:11:38	U	SMPL
PB116313BS - 1	LCSW	13871	3.4065	-7470A	PPB		11 Jan 2019 12:13:45	U	SMPL
CCV19 - 1	CCV19	20686	5.0773	-7470A	PPB		11 Jan 2019 12:15:53	U	SMPL
CCB19 - 1	CCB19	-52	-0.0069	-7470A	PPB		11 Jan 2019 12:18:00	U	SMPL
K1086-02 - 1	CF-1819-W	21	0.0110	-7470A	PPB		11 Jan 2019 12:20:07	U	SMPL
K1086-02DUP - 1	CF-1819-WDUP	9	0.0080	-7470A	PPB		11 Jan 2019 12:22:14	U	SMPL
K1086-02MS - 1	CF-1819-WMS	13943	3.4242	-7470A	PPB		11 Jan 2019 12:24:22	U	SMPL
K1086-02MSD - 1	CF-1819-WMSD	14288	3.5088	-7470A	PPB		11 Jan 2019 12:26:29	U	SMPL
K1100-10 - 1	WATER-DRUMS-A	-8	0.0039	-7470A	PPB		11 Jan 2019 12:28:36	U	SMPL
PB116314BL - 1	PBW	4	0.0068	-7470A	PPB		11 Jan 2019 12:30:45	U	SMPL
PB116314BS - 1	LCSW	13761	3.3795	-7470A	PPB		11 Jan 2019 12:32:53	U	SMPL
K1086-01 - 1	CF-1819-W	-53	-0.0072	-7470A	PPB		11 Jan 2019 12:35:01	U	SMPL
K1100-09 - 1	WATER-DRUMS-A	367	0.0958	-7470A	PPB		11 Jan 2019 12:37:10	U	SMPL
K1100-09DUP - 1	WATER-DRUMS-ADUP	332	0.0872	-7470A	PPB		11 Jan 2019 12:39:18	U	SMPL
CCV20 - 1	CCV20	22137	5.4331	-7470A	PPB		11 Jan 2019 12:41:25	U	SMPL
CCB20 - 1	CCB20	-42	-0.0045	-7470A	PPB		11 Jan 2019 12:43:32	U	SMPL
K1100-09MS - 1	WATER-DRUMS-AMS	17750	4.3575	-7470A	PPB		11 Jan 2019 12:45:40	U	SMPL
K1100-09MSD - 1	WATER-DRUMS-AMSD	17562	4.3114	-7470A	PPB		11 Jan 2019 12:47:47	U	SMPL
K1005-02LX5 - 1		-43	-0.0047	-7470A	PPB		11 Jan 2019 12:49:54	U	SMPL
K1005-02A - 1		16321	4.0072	-7470A	PPB		11 Jan 2019 12:52:02	U	SMPL
K1086-02LX5 - 1		-38	-0.0035	-7470A	PPB		11 Jan 2019 12:54:10	U	SMPL
K1086-02A - 1		16308	4.0040	-7470A	PPB		11 Jan 2019 12:56:17	U	SMPL
K1100-09LX5 - 1		18	0.0102	-7470A	PPB		11 Jan 2019 12:58:26	U	SMPL
K1100-09A - 1		18445	4.5279	-7470A	PPB		11 Jan 2019 13:00:34	U	SMPL
PB116312TB - 1	PB116274TB	0	0.0058	-7470A	PPB		11 Jan 2019 13:02:42	U	SMPL
PB116313TB - 1	PB116277TB	20	0.0107	-7470A	PPB		11 Jan 2019 13:04:50	U	SMPL
CCV21 - 1	CCV21	20751	5.0933	-7470A	PPB		11 Jan 2019 13:06:57	U	SMPL
CCB21 - 1	CCB21	-38	-0.0035	-7470A	PPB		11 Jan 2019 13:09:05	U	SMPL
IDOC1 - 1	IDOC1	15521	3.8110	-7470A	PPB		11 Jan 2019 13:11:12	U	SMPL
IDOC2 - 1	IDOC2	15353	3.7699	-7470A	PPB		11 Jan 2019 13:13:20	U	SMPL
IDOC3 - 1	IDOC3	15523	3.8115	-7470A	PPB		11 Jan 2019 13:15:27	U	SMPL
IDOC4 - 1	IDOC3	15588	3.8275	-7470A	PPB		11 Jan 2019 13:17:35	U	SMPL
CCV22 - 1	CCV22	20525	5.0379	-7470A	PPB		11 Jan 2019 13:19:42	U	SMPL
CCB22 - 1	CCB22	-51	-0.0067	-7470A	PPB		11 Jan 2019 13:21:50	U	SMPL