


LB100271

INSTRUMENT ID: CV1

7471B

 μ Abs.:

59159



Linear

A= 0.0000e+000

B= 1.6930e-004

C= -4.2002e-002

Rho= 0.9998208

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.039	-0.039	15	0.000	15				
0.2	0.200	0.162	-0.038	1204	0.0 %	1204				
2.5	2.500	2.539	0.039	15244	0.0 %	15244				
5.0	5.000	5.137	0.137	30594	0.0 %	30594				
7.5	7.500	7.428	-0.072	44125	0.0 %	44125				
10.0	10.000	9.973	-0.027	59159	0.0 %	59159				

LB100271 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 11 Jan 2019 14:59:15

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	15	-	0.0000	7471B	PPB	11 Jan 2019 15:02:24	S	Std
0.2 - 1	50.2	1204	-	0.2000	7471B	PPB	11 Jan 2019 15:04:43	S	Std
2.5 - 1	52.5	15244	-	2.5000	7471B	PPB	11 Jan 2019 15:07:01	S	Std
5.0 - 1	55.0	30594	-	5.0000	7471B	PPB	11 Jan 2019 15:09:19	S	Std
7.5 - 1	57.5	44125	-	7.5000	7471B	PPB	11 Jan 2019 15:11:36	S	Std
10.0 - 1	510.0	59159	-	10.0000	7471B	PPB	11 Jan 2019 15:21:13	S	Std
ICV58 - 1	ICV58	22334	3.7390	-	7471B	PPB	11 Jan 2019 15:24:00	U	SMPL
ICB58 - 1	ICB58	-101	-0.0591	-	7471B	PPB	11 Jan 2019 15:26:15	U	SMPL
CCV23 - 1	CCV23	29852	5.0118	-	7471B	PPB	11 Jan 2019 15:28:31	U	SMPL
CCB23 - 1	CCB23	-118	-0.0620	-	7471B	PPB	11 Jan 2019 15:30:47	U	SMPL
CRA - 1	CRA	1161	0.1546	-	7471B	PPB	11 Jan 2019 15:33:04	U	SMPL
HighStd - 1	HighStd	58959	9.9395	-	7471B	PPB	11 Jan 2019 15:35:20	U	SMPL
ChkStd - 1	ChkStd	41698	7.0173	-	7471B	PPB	11 Jan 2019 15:37:36	U	SMPL
PB116339BL - 1	PBS	-101	-0.0591	-	7471B	PPB	11 Jan 2019 15:39:55	U	SMPL
PB116339BS - 1	LCSS	20706	3.4634	-	7471B	PPB	11 Jan 2019 15:42:13	U	SMPL
K1011-01 - 1	OR-02-011019-A	1713	0.2480	-	7471B	PPB	11 Jan 2019 15:44:30	U	SMPL
K1011-03 - 1	OR-02-011019-B	1472	0.2072	-	7471B	PPB	11 Jan 2019 15:46:47	U	SMPL
K1113-01 - 1	BCB-COMPOSITE	731	0.0818	-	7471B	PPB	11 Jan 2019 15:49:05	U	SMPL
K1113-01DUP - 1	BCB-COMPOSITEDUP	594	0.0586	-	7471B	PPB	11 Jan 2019 15:51:22	U	SMPL
K1113-01MS - 1	BCB-COMPOSITEMS	22727	3.8056	-	7471B	PPB	11 Jan 2019 15:53:39	U	SMPL
CCV24 - 1	CCV24	29852	5.0118	-	7471B	PPB	11 Jan 2019 15:55:54	U	SMPL
CCB24 - 1	CCB24	-101	-0.0591	-	7471B	PPB	11 Jan 2019 15:58:12	U	SMPL
K1113-01MSD - 1	BCB-COMPOSITMSD	21855	3.6579	-	7471B	PPB	11 Jan 2019 16:00:29	U	SMPL
K1118-01 - 1	FENCE-SOIL-PILE-1	5828	0.9447	-	7471B	PPB	11 Jan 2019 16:02:45	U	SMPL
K1118-05 - 1	TESTPIT-SOIL-1	4086	0.6497	-	7471B	PPB	11 Jan 2019 16:05:02	U	SMPL
K1119-01 - 1	DS-A	8985	1.4791	-	7471B	PPB	11 Jan 2019 16:07:19	U	SMPL
K1119-03 - 1	DS-B	4502	0.7202	-	7471B	PPB	11 Jan 2019 16:09:35	U	SMPL
K1119-05 - 1	DS-C	7661	1.2550	-	7471B	PPB	11 Jan 2019 16:11:52	U	SMPL
K1119-07 - 1	DS-D	6264	1.0185	-	7471B	PPB	11 Jan 2019 16:14:09	U	SMPL
K1124-01 - 1	ER-SOIL	2353	0.3564	-	7471B	PPB	11 Jan 2019 16:16:27	U	SMPL
K1113-01LX5 - 1		148	-0.0169	-	7471B	PPB	11 Jan 2019 16:18:45	U	SMPL
K1113-01A - 1		22806	3.8189	-	7471B	PPB	11 Jan 2019 16:26:28	U	SMPL
CCV25 - 1	CCV25	29811	5.0049	-	7471B	PPB	11 Jan 2019 16:28:44	U	SMPL
CCB25 - 1	CCB25	-94	-0.0579	-	7471B	PPB	11 Jan 2019 16:31:01	U	SMPL