



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100273

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	01/11/2019	16:05
2	CAL2	1	Water	NA	NA	21.1	7.01	01/11/2019	16:06
3	CAL3	1	Water	NA	NA	21.1	10.02	01/11/2019	16:07
4	ICV	1	Water	NA	NA	21.1	7.00	01/11/2019	16:08
5	CCV1	1	Water	NA	NA	21.1	2.01	01/11/2019	16:09
6	K1127-03	1	Water	NA	NA	18.2	5.88	01/11/2019	16:10
7	K1127-03DUP	1	Water	NA	NA	18.3	5.89	01/11/2019	16:11
8	CCV2	1	Water	NA	NA	21.2	12.02	01/11/2019	16:12

LB100273

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity K1127

WorkList ID : 121251

Date : 1/11/2019 3:56:33 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/11/2019	Water	K1127-03~	Corrosivity	Cool 4 deg C	PSEG01	E61	COMP-1	01/11/2019	9040C

Date/Time 01-11-19 4:00 PM
Received by: Jb
Relinquished by: Jb

Date/Time 01-11-19 5:30 PM
Received by: Jb
Relinquished by: Jb