


LB100287

INSTRUMENT ID: CV1

7470A

 μ Abs.:

42899



Linear

A= 0.0000e+000

B= 2.3442e-004

C= 6.0105e-002

Rho= 0.9996778

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.051	0.051	-40	0.000	-40				
0.2	0.200	0.229	0.029	720	0.0 %	720				
2.5	2.500	2.428	-0.072	10101	0.0 %	10101				
5.0	5.000	5.045	0.045	21265	0.0 %	21265				
7.5	7.500	7.331	-0.169	31018	0.0 %	31018				
10.0	10.000	10.116	0.116	42899	0.0 %	42899	0			

LB100287 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Jan 2019 12:07:09

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-40	-	0.0000	7470A	PPB	14 Jan 2019 12:07:44	S	Std
0.2 - 1	50.2	720	-	0.2000	7470A	PPB	14 Jan 2019 12:09:53	S	Std
2.5 - 1	52.5	10101	-	2.5000	7470A	PPB	14 Jan 2019 12:12:02	S	Std
5.0 - 1	55.0	21265	-	5.0000	7470A	PPB	14 Jan 2019 12:14:10	S	Std
7.5 - 1	57.5	31018	-	7.5000	7470A	PPB	14 Jan 2019 12:16:18	S	Std
10.0 - 1	510.0	42899	-	10.0000	7470A	PPB	14 Jan 2019 12:21:59	S	Std
ICV60 - 1	ICV60	16776	3.9927	-	7470A	PPB	14 Jan 2019 12:26:59	U	SMPL
ICB60 - 1	ICB60	-89	0.0392	-	7470A	PPB	14 Jan 2019 12:29:05	U	SMPL
CCV30 - 1	CCV30	19869	4.7177	-	7470A	PPB	14 Jan 2019 12:31:12	U	SMPL
CCB30 - 1	CCB30	-106	0.0353	-	7470A	PPB	14 Jan 2019 12:33:19	U	SMPL
CRA - 1	CRA	726	0.2303	-	7470A	PPB	14 Jan 2019 12:35:26	U	SMPL
HighStd - 1	HighStd	42334	9.9839	-	7470A	PPB	14 Jan 2019 12:37:33	U	SMPL
ChkStd - 1	ChkStd	29198	6.9046	-	7470A	PPB	14 Jan 2019 12:39:40	U	SMPL
PB116340BL - 1	PBW	-132	0.0292	-	7470A	PPB	14 Jan 2019 12:41:48	U	SMPL
PB116340BS - 1	LCSW	16816	4.0020	-	7470A	PPB	14 Jan 2019 12:43:56	U	SMPL
K1106-01 - 1	TS-A	11	0.0627	-	7470A	PPB	14 Jan 2019 12:52:32	U	SMPL
K1106-01DUP - 1	TS-ADUP	-29	0.0533	-	7470A	PPB	14 Jan 2019 12:54:40	U	SMPL
K1106-01MS - 1	TS-AMS	16672	3.9683	-	7470A	PPB	14 Jan 2019 12:56:48	U	SMPL
K1106-01MSD - 1	TS-AMSD	15748	3.7517	-	7470A	PPB	14 Jan 2019 12:58:57	U	SMPL
K1106-04 - 1	TS-D	-80	0.0414	-	7470A	PPB	14 Jan 2019 13:01:04	U	SMPL
CCV31 - 1	CCV31	19832	4.7090	-	7470A	PPB	14 Jan 2019 13:03:12	U	SMPL
CCB31 - 1	CCB31	-107	0.0350	-	7470A	PPB	14 Jan 2019 13:05:19	U	SMPL
K1106-08 - 1	TS-H	4	0.0610	-	7470A	PPB	14 Jan 2019 13:07:25	U	SMPL
K1106-10 - 1	TS-J	-1	0.0599	-	7470A	PPB	14 Jan 2019 13:09:32	U	SMPL
PB116340TB - 1	PB116315TB	-103	0.0360	-	7470A	PPB	14 Jan 2019 13:11:39	U	SMPL
K1106-01LX5 - 1		-41	0.0505	-	7470A	PPB	14 Jan 2019 13:13:46	U	SMPL
K1106-01A - 1		15482	3.6893	-	7470A	PPB	14 Jan 2019 13:15:53	U	SMPL
CCV32 - 1	CCV32	20762	4.9270	-	7470A	PPB	14 Jan 2019 13:18:01	U	SMPL
CCB32 - 1	CCB32	-111	0.0341	-	7470A	PPB	14 Jan 2019 13:20:09	U	SMPL