

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB100288

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP70044
calibration std. phosphate 1 ppm	WP70895
calibration std. phosphate 0.5 ppm	WP70894
calibration std. phosphate 0.3 ppm	WP70893
calibration std. 0 ppm	WP70890
phosphate ICV std.	WP70897
phosphate CCV std.	WP70896
calibration std. phosphate 0.1 ppm	WP70892
calibration std. phosphate 0.05 ppm	WP70891
Combined reagent	WP70899
11 N sulfuric acid	WP66905
Phenolphthalein indicator	WP66907
Sodium hydroxide, 1N	WP66924

Intercept: -0.0008Slope: 0.651Regression: 0.999827

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	01/14/2019	11:13
2	CAL2	0.05	1	50	50	0.033	01/14/2019	11:13
3	CAL3	0.10	1	50	50	0.066	01/14/2019	11:14
4	CAL4	0.30	1	50	50	0.186	01/14/2019	11:14
5	CAL5	0.50	1	50	50	0.330	01/14/2019	11:15
6	CAL6	1.00	1	50	50	0.650	01/14/2019	11:16



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB100288

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.325	0.500	01/14/2019	11:16
2	ICB		1	50	50	0.000	0.001	01/14/2019	11:16
3	CCV1	0.50	1	50	50	0.327	0.504	01/14/2019	11:17
4	CCB1		1	50	50	0.000	0.001	01/14/2019	11:18
5	PB116369BL		1	50	50	0.000	0.001	01/14/2019	11:18
6	PB116369BS	0.50	1	50	50	0.365	0.562	01/14/2019	11:19
7	K1085-01		1	50	50	0.028	0.044	01/14/2019	11:19
8	K1085-01DUP		1	50	50	0.027	0.043	01/14/2019	11:20
9	K1085-01MS	0.50	1	50	50	0.353	0.543	01/14/2019	11:21
10	K1085-01MSD	0.50	1	50	50	0.352	0.542	01/14/2019	11:21
11	CCV2	0.50	1	50	50	0.328	0.505	01/14/2019	11:22
12	CCB2		1	50	50	0.000	0.001	01/14/2019	11:22