



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100289

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	01/14/2019	13:15
2	CAL2	1	Water	NA	NA	21.1	7.01	01/14/2019	13:16
3	CAL3	1	Water	NA	NA	21.1	10.02	01/14/2019	13:17
4	ICV	1	Water	NA	NA	21.1	7.00	01/14/2019	13:18
5	CCV1	1	Water	NA	NA	21.2	2.01	01/14/2019	13:19
6	K1139-22	1	Water	NA	NA	14.1	6.53	01/14/2019	13:20
7	K1139-22DUP	1	Water	NA	NA	14.2	6.54	01/14/2019	13:21
8	CCV2	1	Water	NA	NA	21.2	12.02	01/14/2019	13:22

LIB 100289

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosion k1139

WorkList ID : 121287

Date : 1/14/2019 1:05:06 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/12/2019	Water	K1139-22	Corrosivity	Cool 4 deg C	PSEG01	E61	T-CONTAINMENT	01/12/2019	9040C

Date/Time 01-14-19 1:10 PM
Received by: JO
Relinquished by: [Signature]

Date/Time 01-14-19 2:00 PM
Received by: [Signature]
Relinquished by: JO