



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 1/15/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:00
In Date: 01/14/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 01/15/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1

QC:LB100294

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K1007-01	NB-07-011419	29	1.1	9.66	9.08	93.2
K1135-01	CPSB-19(10-15)	1	1.15	9.96	8.35	81.7
K1135-02	CPSB-19(15-20)	2	1.18	9.69	3.97	32.8
K1135-03	CPSB-12(5-6)	3	1.19	9.97	9.06	89.6
K1135-04	CPSB-12(10-15)	4	1.15	9.97	8.68	85.4
K1139-01	S-1	5	1.13	9.58	7.49	75.3
K1139-02	S-1	6	1.13	9.58	7.49	75.3
K1139-03	S-2	7	1.17	9.71	7.71	76.6
K1139-04	S-2	8	1.17	9.71	7.71	76.6
K1139-05	S-3	9	1.11	9.91	7.17	68.9
K1139-06	S-3	10	1.11	9.91	7.17	68.9
K1139-07	S-4	11	1.19	9.82	7.38	71.7
K1139-08	S-4	12	1.19	9.82	7.38	71.7
K1139-09	S-5	13	1.18	9.67	7.49	74.3
K1139-10	S-5	14	1.18	9.67	7.49	74.3
K1139-11	S-6	15	1.12	9.64	7.1	70.2
K1139-12	S-6	16	1.12	9.64	7.1	70.2
K1139-13	S-7	17	1.17	9.66	7.04	69.1
K1139-14	S-7	18	1.17	9.66	7.04	69.1
K1139-15	S-8	19	1.13	9.76	7.32	71.7
K1139-16	S-8	20	1.13	9.76	7.32	71.7
K1139-17	S-9	21	1.1	9.59	7.31	73.1
K1139-18	S-9	22	1.1	9.59	7.31	73.1
K1139-19	CONCRETE	23	1.00	2.00	2.00	100.0
K1142-01	FK-BPM-01-D9-E9-D9A-E9A	24	1.14	9.91	9.23	92.2
K1142-02	FK-BPM-01-D9-E9-D9A-E9A	25	1.16	9.79	9.09	91.9
K1144-01	AB	26	1.00	2.00	2.00	100.0
K1144-02	CD	27	1.00	2.00	2.00	100.0
K1144-03	OIL	28	1.00	2.00	2.00	100.0

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$