

LB100305


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

54497



A= 0.0000e+000

B= 1.8406e-004

C= -5.3239e-003

Rho= 0.9999758

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.009	0.009	77	0.000	77				
0.2	0.200	0.214	0.014	1192	0.0 %	1192				
2.5	2.500	2.468	-0.032	13439	0.0 %	13439				
5.0	5.000	5.023	0.023	27318	0.0 %	27318				
7.5	7.500	7.461	-0.039	40566	0.0 %	40566				
10.0	10.000	10.025	0.025	54497	0.0 %	54497				

LB100305 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 Jan 2019 09:31:32

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	77	-	0.0007471B	PPB		15 Jan 2019 09:33:24	S	Std
0.2 - 1	50.2	1192	-	0.20007471B	PPB		15 Jan 2019 09:35:43	S	Std
2.5 - 1	52.5	13439	-	2.50007471B	PPB		15 Jan 2019 09:38:01	S	Std
5.0 - 1	55.0	27318	-	5.00007471B	PPB		15 Jan 2019 09:40:19	S	Std
7.5 - 1	57.5	40566	-	7.50007471B	PPB		15 Jan 2019 09:42:36	S	Std
10.0 - 1	510.0	54497	-	10.00007471B	PPB		15 Jan 2019 09:44:55	S	Std
ICV62 - 1	ICV62	22505	4.1368	-7471B	PPB		15 Jan 2019 10:01:38	U	SMPL
ICB62 - 1	ICB62	-47	-0.0140	-7471B	PPB		15 Jan 2019 10:03:53	U	SMPL
CCV36 - 1	CCV36	27346	5.0278	-7471B	PPB		15 Jan 2019 10:06:11	U	SMPL
CCB36 - 1	CCB36	-63	-0.0169	-7471B	PPB		15 Jan 2019 10:08:27	U	SMPL
CRA - 1	CRA	1180	0.2119	-7471B	PPB		15 Jan 2019 10:10:45	U	SMPL
HighStd - 1	HighStd	54827	10.0859	-7471B	PPB		15 Jan 2019 10:13:00	U	SMPL
ChkStd - 1	ChkStd	39388	7.2442	-7471B	PPB		15 Jan 2019 10:15:17	U	SMPL
PB116379BL - 1	PBS	-18	-0.0086	-7471B	PPB		15 Jan 2019 10:17:35	U	SMPL
PB116379BS - 1	LCSS	19282	3.5436	-7471B	PPB		15 Jan 2019 10:19:54	U	SMPL
K1015-05 - 1	SU-02-011119-A	2654	0.4832	-7471B	PPB		15 Jan 2019 10:22:11	U	SMPL
K1015-05DUP - 1	SU-02-011119-ADUP	2599	0.4730	-7471B	PPB		15 Jan 2019 10:24:29	U	SMPL
K1015-05MS - 1	SU-02-011119-AMS	20867	3.8354	-7471B	PPB		15 Jan 2019 10:26:47	U	SMPL
K1015-05MSD - 1	SU-02-011119-AMSD	20702	3.8050	-7471B	PPB		15 Jan 2019 10:29:04	U	SMPL
K1015-07 - 1	SU-02-011119-B	2074	0.3764	-7471B	PPB		15 Jan 2019 10:31:22	U	SMPL
CCV37 - 1	CCV37	27706	5.0941	-7471B	PPB		15 Jan 2019 10:33:39	U	SMPL
CCB37 - 1	CCB37	-82	-0.0204	-7471B	PPB		15 Jan 2019 10:35:55	U	SMPL
K1126-01 - 1	RT-3451	1612	0.2914	-7471B	PPB		15 Jan 2019 10:38:12	U	SMPL
K1128-01 - 1	RBR-200053	2075	0.3766	-7471B	PPB		15 Jan 2019 10:40:28	U	SMPL
K1128-03 - 1	RBR-200036	1962	0.3558	-7471B	PPB		15 Jan 2019 10:42:44	U	SMPL
K1129-01 - 1	P-1	1897	0.3438	-7471B	PPB		15 Jan 2019 10:45:01	U	SMPL
K1129-03 - 1	P-2	1069	0.1914	-7471B	PPB		15 Jan 2019 10:47:18	U	SMPL
K1131-01 - 1	DGA-01-011019	328	0.0550	-7471B	PPB		15 Jan 2019 10:49:35	U	SMPL
K1131-02 - 1	57-01-011019	226	0.0363	-7471B	PPB		15 Jan 2019 10:51:53	U	SMPL
K1131-03 - 1	QD-01-011019	267	0.0438	-7471B	PPB		15 Jan 2019 10:54:11	U	SMPL
K1139-01 - 1	S-1	8825	1.6190	-7471B	PPB		15 Jan 2019 10:56:29	U	SMPL
K1139-03 - 1	S-2	43047	7.9177	-7471B	PPB		15 Jan 2019 10:58:45	U	SMPL
CCV38 - 1	CCV38	27596	5.0739	-7471B	PPB		15 Jan 2019 11:01:02	U	SMPL
CCB38 - 1	CCB38	-75	-0.0191	-7471B	PPB		15 Jan 2019 11:03:20	U	SMPL
K1139-05 - 1	S-3	4159	0.7602	-7471B	PPB		15 Jan 2019 11:05:38	U	SMPL
K1139-07 - 1	S-4	80434	14.7990	-7471B	PPB		15 Jan 2019 11:07:55	U	SMPL
K1139-09 - 1	S-5	24282	4.4639	-7471B	PPB		15 Jan 2019 11:10:11	U	SMPL
K1139-11 - 1	S-6	9374	1.7200	-7471B	PPB		15 Jan 2019 11:12:31	U	SMPL
K1139-13 - 1	S-7	18121	3.3299	-7471B	PPB		15 Jan 2019 11:14:48	U	SMPL
K1139-15 - 1	S-8	15122	2.7780	-7471B	PPB		15 Jan 2019 11:17:06	U	SMPL
K1139-17 - 1	S-9	7817	1.4334	-7471B	PPB		15 Jan 2019 11:19:23	U	SMPL
K1142-01 - 1	FK-BPM-01-D9-E9-D9A-E9A	2358	0.4287	-7471B	PPB		15 Jan 2019 11:21:42	U	SMPL
K1145-03 - 1	390110055	131	0.0188	-7471B	PPB		15 Jan 2019 11:24:00	U	SMPL
CCV39 - 1	CCV39	28168	5.1791	-7471B	PPB		15 Jan 2019 11:26:16	U	SMPL
CCB39 - 1	CCB39	-40	-0.0127	-7471B	PPB		15 Jan 2019 11:28:33	U	SMPL
K1015-05LX5 - 1		539	0.0939	-7471B	PPB		15 Jan 2019 11:30:51	U	SMPL
K1015-05A - 1		25331	4.6570	-7471B	PPB		15 Jan 2019 11:33:07	U	SMPL
K1139-07DLX2 - 1		40822	7.5082	-7471B	PPB		15 Jan 2019 11:35:24	U	SMPL
CCV40 - 1	CCV40	28177	5.1808	-7471B	PPB		15 Jan 2019 11:37:42	U	SMPL
CCB40 - 1	CCB40	-32	-0.0112	-7471B	PPB		15 Jan 2019 11:40:01	U	SMPL