



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 01/15/2019

TEMP1 IN: 103 01/15/2019 07:00 TEMP OUT: 104 01/15/2019 08:05

TEMP2 IN: 104 01/15/2019 09:00 TEMP OUT: 104 01/15/2019 10:05

TEMP3 IN: 105 01/15/2019 11:00 TEMP OUT: 105 01/15/2019 12:40

TEMP4 IN: 105 01/15/2019 13:20 TEMP OUT: 105 01/15/2019 15:00

Run Number: LB100308

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB100308BL	LB100308BL	1.4012	1.4012	100	1.4012	1.4012	1.4012	0.0000	0
2	LB100308BS	LB100308BS	1.4062	1.4062	100	1.4563	1.4563	1.4563	0.0501	501
3	K1127-02	COMP-1	1.4040	1.4040	30	1.4220	1.4220	1.4220	0.0180	600
4	K1139-21	T-CONTAINMENT	1.4012	1.4012	300	1.4078	1.4078	1.4078	0.0066	22
5	K1140-01	001M-WILLETS-PT-BLVD (DEC)	1.3874	1.3874	250	1.4054	1.4054	1.4054	0.0180	72
6	K1141-01	001A	1.3970	1.3970	200	1.4055	1.4055	1.4055	0.0085	42.5
7	K1141-02	002A	1.4064	1.4064	150	1.4152	1.4152	1.4152	0.0088	58.7
8	K1141-02DUP	002ADUP	1.4011	1.4011	150	1.4100	1.4100	1.4100	0.0089	59.3

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS011519

WorkList ID : 121298


Date : 1/15/2019 7:31:16 AM

20100308

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/16/2019	Water	K1127-02	F TSS	Cool 4 deg C	PSEG01	E61	COMP-1	01/11/2019	SM2540 D
01/19/2019	Water	K1139-21	G TSS	Cool 4 deg C	PSEG01	E61	T-CONTAINMENT	01/12/2019	SM2540 D
01/16/2019	Water	K1140-01	B TSS	Cool 4 deg C	TULL01	F12	001M-WILLETTS-PT-BLVD(DE	01/14/2019	SM2540 D
01/16/2019	Water	K1141-01	✓ TSS	Cool 4 deg C	WILL06	K12	001A	01/14/2019	SM2540 D
01/16/2019	Water	K1141-02	✓ TSS	Cool 4 deg C	WILL06	K12	002A	01/14/2019	SM2540 D

FP-12791048

Date/Time 01/15/19, 08:00
 Received by: 
 Relinquished by: 

Date/Time 01/15/19, 12:00
 Received by: 
 Relinquished by: 