


LB100310
INSTRUMENT ID : CVI

7470A

μ Abs. :

41325



Linear

A= 0.0000e+000

B= 2.4357e-004

C= -1.2651e-002

Rho= 0.9997688

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.018	0.018	125	0.000	125				
0.2	0.200	0.158	-0.042	699	0.0 %	699				
2.5	2.500	2.518	0.018	10390	0.0 %	10390				
5.0	5.000	5.102	0.102	21000	0.0 %	21000				
7.5	7.500	7.351	-0.149	30234	0.0 %	30234				
10.0	10.000	10.053	0.053	41325	0.0 %	41325	0			

LB100310 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 15 Jan 2019 12:25:50				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50.0	125	-	0.0000	7470A	PPB	15 Jan 2019 12:27:34	S Std
0.2 - 1	50.2	699	-	0.2000	7470A	PPB	15 Jan 2019 12:29:43	S Std
2.5 - 1	52.5	10390	-	2.5000	7470A	PPB	15 Jan 2019 12:31:52	S Std
5.0 - 1	55.0	21000	-	5.0000	7470A	PPB	15 Jan 2019 12:34:01	S Std
7.5 - 1	57.5	30234	-	7.5000	7470A	PPB	15 Jan 2019 12:36:09	S Std
10.0 - 1	610.0	41325	-	10.0000	7470A	PPB	15 Jan 2019 12:42:44	S Std
ICV63 - 1	ICV63	15796	3.8348	-	7470A	PPB	15 Jan 2019 12:46:42	U SMPL
ICB63 - 1	ICB63	19	-0.0080	-	7470A	PPB	15 Jan 2019 12:48:49	U SMPL
CCV41 - 1	CCV41	19913	4.8375	-	7470A	PPB	15 Jan 2019 12:50:56	U SMPL
CCB41 - 1	CCB41	10	-0.0102	-	7470A	PPB	15 Jan 2019 12:53:03	U SMPL
CRA - 1	CRA	835	0.1907	-	7470A	PPB	15 Jan 2019 12:55:09	U SMPL
HighStd - 1	HighStd	41235	10.0309	-	7470A	PPB	15 Jan 2019 12:57:16	U SMPL
ChkStd - 1	ChkStd	28835	7.0107	-	7470A	PPB	15 Jan 2019 12:59:23	U SMPL
PB116380BL - 1	PBW	-46	-0.0239	-	7470A	PPB	15 Jan 2019 13:01:31	U SMPL
PB116380BS - 1	LCSW	15644	3.7977	-	7470A	PPB	15 Jan 2019 13:03:40	U SMPL
K1127-03 - 1	COMP-1	112	0.0146	-	7470A	PPB	15 Jan 2019 13:17:12	U SMPL
K1127-03DUP - 1	COMP-1DUP	73	0.0051	-	7470A	PPB	15 Jan 2019 13:19:20	U SMPL
K1127-03MS - 1	COMP-1MS	15813	3.8389	-	7470A	PPB	15 Jan 2019 13:21:28	U SMPL
K1127-03MSD - 1	COMP-1MSD	16419	3.9865	-	7470A	PPB	15 Jan 2019 13:23:37	U SMPL
PB116381BL - 1	PBW	-61	-0.0275	-	7470A	PPB	15 Jan 2019 13:25:44	U SMPL
CCV42 - 1	CCV42	19644	4.7720	-	7470A	PPB	15 Jan 2019 13:27:51	U SMPL
CCB42 - 1	CCB42	23	-0.0070	-	7470A	PPB	15 Jan 2019 13:29:57	U SMPL
PB116381BS - 1	LCSW	15644	3.7977	-	7470A	PPB	15 Jan 2019 13:32:04	U SMPL
K1015-06 - 1	SU-02-011119-A	10	-0.0102	-	7470A	PPB	15 Jan 2019 13:34:11	U SMPL
K1015-06DUP - 1	SU-02-011119-ADUP	27	-0.0061	-	7470A	PPB	15 Jan 2019 13:36:17	U SMPL
K1015-06MS - 1	SU-02-011119-AMS	15821	3.8409	-	7470A	PPB	15 Jan 2019 13:38:25	U SMPL
K1015-06MSD - 1	SU-02-011119-AMSD	15724	3.8172	-	7470A	PPB	15 Jan 2019 13:40:32	U SMPL
K1015-08 - 1	SU-02-011119-B	0	-0.0127	-	7470A	PPB	15 Jan 2019 13:42:40	U SMPL
K1126-02 - 1	RT-3451	42	-0.0024	-	7470A	PPB	15 Jan 2019 13:44:48	U SMPL
K1128-02 - 1	RBR-200053	45	-0.0017	-	7470A	PPB	15 Jan 2019 13:46:57	U SMPL
K1128-04 - 1	RBR-200036	28	-0.0058	-	7470A	PPB	15 Jan 2019 13:49:06	U SMPL
CCV43 - 1	CCV43	19728	4.7925	-	7470A	PPB	15 Jan 2019 13:51:13	U SMPL
CCB43 - 1	CCB43	64	0.0029	-	7470A	PPB	15 Jan 2019 13:53:20	U SMPL
PB116380TB - 1	PB116343TB	121	0.0168	-	7470A	PPB	15 Jan 2019 13:55:28	U SMPL
PB116381TB - 1	PB116341TB	65	0.0032	-	7470A	PPB	15 Jan 2019 13:57:35	U SMPL
K1127-03LX5 - 1		125	0.0178	-	7470A	PPB	15 Jan 2019 13:59:42	U SMPL
K1127-03A - 1		15515	3.7663	-	7470A	PPB	15 Jan 2019 14:01:49	U SMPL
K1015-06LX5 - 1		-43	-0.0231	-	7470A	PPB	15 Jan 2019 14:03:56	U SMPL
K1015-06A - 1		16086	3.9054	-	7470A	PPB	15 Jan 2019 14:06:04	U SMPL
CCV44 - 1	CCV44	19803	4.8107	-	7470A	PPB	15 Jan 2019 14:08:11	U SMPL
CCB44 - 1	CCB44	67	0.0037	-	7470A	PPB	15 Jan 2019 14:10:19	U SMPL