



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: AMANDEEP

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB100314

Ambient Barometric Pressure (mmHg): 750

Barometric Scale ID: 051106

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	81.2	01/15/2019	13:15
2	K1139-22		1	8	>100.0	>212.0	>212.0	01/15/2019	13:24
3	K1139-22DUP		1	8	>100.0	>212.0	>212.0	01/15/2019	13:32

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

UB100314

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP011519

WorkList ID : 121315

Date : 1/15/2019 9:18:32 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/21/2019	Water	K1139-22	Flash Point	Cool 4 deg C	PSEG01	E61	T-CONTAINMENT	01/12/2019	1010A

Date/Time 01-15-19 13:00
 Received by: AO
 Relinquished by: AO

Date/Time 01-18-19 13:35
 Received by: AO
 Relinquished by: AO