

LB100325


7470A

INSTRUMENT ID : CV1

Linear

 μ Abs. :

40743



A= 0.0000e+000

B= 2.4781e-004

C= 1.7768e-002

Rho= 0.9997875

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.015	0.015	-10	0.000	-10				
0.2	0.200	0.229	0.029	852	0.0 %	852				
2.5	2.500	2.529	0.029	10133	0.0 %	10133				
5.0	5.000	4.940	-0.060	19864	0.0 %	19864				
7.5	7.500	7.373	-0.127	29680	0.0 %	29680				
10.0	10.000	10.114	0.114	40743	0.0 %	40743	0			

LB100325 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Jan 2019 09:22:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	-10	-	0.0000	7470A	PPB	16 Jan 2019 09:26:59	S	Std
0.2 - 1	50-2	852	-	0.2000	7470A	PPB	16 Jan 2019 09:29:08	S	Std
2.5 - 1	52-5	10133	-	2.5000	7470A	PPB	16 Jan 2019 09:31:17	S	Std
5.0 - 1	55-0	19864	-	5.0000	7470A	PPB	16 Jan 2019 09:33:26	S	Std
7.5 - 1	57-5	29680	-	7.5000	7470A	PPB	16 Jan 2019 09:35:34	S	Std
10.0 - 1	510-0	40743	-	10.0000	7470A	PPB	16 Jan 2019 09:37:42	S	Std
ICV64 - 1	ICV64	15484	3.8548	-	7470A	PPB	16 Jan 2019 09:42:06	U	SMPL
ICB64 - 1	ICB64	-91	-0.0048	-	7470A	PPB	16 Jan 2019 09:44:13	U	SMPL
CCV45 - 1	CCV45	20342	5.0586	-	7470A	PPB	16 Jan 2019 09:46:19	U	SMPL
CCB45 - 1	CCB45	-153	-0.0201	-	7470A	PPB	16 Jan 2019 09:48:26	U	SMPL
CRA - 1	CRA	711	0.1940	-	7470A	PPB	16 Jan 2019 09:50:33	U	SMPL
HighStd - 1	HighStd	40260	9.9945	-	7470A	PPB	16 Jan 2019 09:52:40	U	SMPL
ChkStd - 1	ChkStd	27583	6.8530	-	7470A	PPB	16 Jan 2019 09:54:47	U	SMPL
PB116400BL - 1	PBW	-175	-0.0256	-	7470A	PPB	16 Jan 2019 09:56:56	U	SMPL
PB116400BS - 1	LCSW	13879	3.4571	-	7470A	PPB	16 Jan 2019 09:59:03	U	SMPL
K1007-02 - 1	NB-07-011419	-114	-0.0105	-	7470A	PPB	16 Jan 2019 10:01:11	U	SMPL
K1007-02DUP - 1	NB-07-011419DUP	-1	0.0175	-	7470A	PPB	16 Jan 2019 10:03:19	U	SMPL
K1007-02MS - 1	NB-07-011419MS	13864	3.4534	-	7470A	PPB	16 Jan 2019 10:05:27	U	SMPL
K1007-02MSD - 1	NB-07-011419MSD	14072	3.5049	-	7470A	PPB	16 Jan 2019 10:07:36	U	SMPL
K1065-02 - 1	STOCKPILE	-89	-0.0043	-	7470A	PPB	16 Jan 2019 10:09:43	U	SMPL
CCV46 - 1	CCV46	20423	5.0787	-	7470A	PPB	16 Jan 2019 10:11:50	U	SMPL
CCB46 - 1	CCB46	-82	-0.0026	-	7470A	PPB	16 Jan 2019 10:13:57	U	SMPL
K1139-02 - 1	S-1	-19	0.0131	-	7470A	PPB	16 Jan 2019 10:16:04	U	SMPL
K1139-04 - 1	S-2	16	0.0217	-	7470A	PPB	16 Jan 2019 10:18:11	U	SMPL
K1139-06 - 1	S-3	26	0.0242	-	7470A	PPB	16 Jan 2019 10:20:19	U	SMPL
K1139-08 - 1	S-4	39	0.0274	-	7470A	PPB	16 Jan 2019 10:22:26	U	SMPL
K1139-10 - 1	S-5	-20	0.0128	-	7470A	PPB	16 Jan 2019 10:24:34	U	SMPL
K1139-12 - 1	S-6	0	0.0178	-	7470A	PPB	16 Jan 2019 10:26:42	U	SMPL
K1139-14 - 1	S-7	7	0.0195	-	7470A	PPB	16 Jan 2019 10:28:50	U	SMPL
K1139-16 - 1	S-8	-3	0.0170	-	7470A	PPB	16 Jan 2019 10:30:59	U	SMPL
K1139-18 - 1	S-9	1	0.0180	-	7470A	PPB	16 Jan 2019 10:33:07	U	SMPL
K1144-01 - 1	AB	-13	0.0145	-	7470A	PPB	16 Jan 2019 10:35:14	U	SMPL
CCV47 - 1	CCV47	20315	5.0520	-	7470A	PPB	16 Jan 2019 10:37:21	U	SMPL
CCB47 - 1	CCB47	-81	-0.0023	-	7470A	PPB	16 Jan 2019 10:39:28	U	SMPL
K1144-02 - 1	CD	28	0.0247	-	7470A	PPB	16 Jan 2019 10:41:36	U	SMPL
K1144-03 - 1	OIL	7	0.0195	-	7470A	PPB	16 Jan 2019 10:43:43	U	SMPL
PB116401BL - 1	PBW	2	0.0183	-	7470A	PPB	16 Jan 2019 10:45:51	U	SMPL
PB116401BS - 1	LCSW	15274	3.8028	-	7470A	PPB	16 Jan 2019 10:47:58	U	SMPL
K1139-22 - 1	T-CONTAINMENT	-36	0.0088	-	7470A	PPB	16 Jan 2019 10:50:06	U	SMPL
K1139-22DUP - 1	T-CONTAINMENTDUP	-4	0.0168	-	7470A	PPB	16 Jan 2019 10:52:14	U	SMPL
K1139-22MS - 1	T-CONTAINMENTMS	15731	3.9160	-	7470A	PPB	16 Jan 2019 10:54:22	U	SMPL
K1139-22MSD - 1	T-CONTAINMENTMSD	14441	3.5963	-	7470A	PPB	16 Jan 2019 10:56:30	U	SMPL
PB116402BL - 1	PBW	-40	0.0079	-	7470A	PPB	16 Jan 2019 10:58:39	U	SMPL
PB116402BS - 1	LCSW	14580	3.6308	-	7470A	PPB	16 Jan 2019 11:00:46	U	SMPL
CCV48 - 1	CCV48	19386	4.8217	-	7470A	PPB	16 Jan 2019 11:02:53	U	SMPL
CCB48 - 1	CCB48	-70	0.0004	-	7470A	PPB	16 Jan 2019 11:05:01	U	SMPL
K1127-02 - 1	COMP-1	456	0.1308	-	7470A	PPB	16 Jan 2019 11:07:08	U	SMPL
K1136-02 - 1	MW-103	-135	-0.0157	-	7470A	PPB	16 Jan 2019 11:09:15	U	SMPL
K1136-05 - 1	EB_2019.01.11	57	0.0319	-	7470A	PPB	16 Jan 2019 11:11:23	U	SMPL
K1139-21 - 1	T-CONTAINMENT	71	0.0354	-	7470A	PPB	16 Jan 2019 11:13:30	U	SMPL
K1139-21DUP - 1	T-CONTAINMENTDUP	52	0.0307	-	7470A	PPB	16 Jan 2019 11:15:38	U	SMPL
K1139-21MS - 1	T-CONTAINMENTMS	18607	4.6287	-	7470A	PPB	16 Jan 2019 11:17:46	U	SMPL
K1139-21MSD - 1	T-CONTAINMENTMSD	16461	4.0969	-	7470A	PPB	16 Jan 2019 11:19:54	U	SMPL
PB116400TB - 1	PB116370TB	-37	0.0086	-	7470A	PPB	16 Jan 2019 11:22:03	U	SMPL
PB116402TB - 1	PB116372TB	10	0.0202	-	7470A	PPB	16 Jan 2019 11:24:11	U	SMPL
CCV49 - 1	CCV49	19539	4.8597	-	7470A	PPB	16 Jan 2019 11:26:19	U	SMPL
CCB49 - 1	CCB49	-56	0.0039	-	7470A	PPB	16 Jan 2019 11:28:26	U	SMPL
K1007-02LX5 - 1		0	0.0178	-	7470A	PPB	16 Jan 2019 11:30:33	U	SMPL
K1007-02A - 1		15868	3.9500	-	7470A	PPB	16 Jan 2019 11:32:40	U	SMPL
K1139-22LX5 - 1		-28	0.0108	-	7470A	PPB	16 Jan 2019 11:34:48	U	SMPL
K1139-22A - 1		16535	4.1153	-	7470A	PPB	16 Jan 2019 11:36:55	U	SMPL
K1139-21LX5 - 1		-34	0.0093	-	7470A	PPB	16 Jan 2019 11:39:02	U	SMPL
K1139-21A - 1		17414	4.3331	-	7470A	PPB	16 Jan 2019 11:45:39	U	SMPL
CCV50 - 1	CCV50	19424	4.8312	-	7470A	PPB	16 Jan 2019 11:47:47	U	SMPL
CCB50 - 1	CCB50	-70	0.0004	-	7470A	PPB	16 Jan 2019 11:49:55	U	SMPL