


LB100332

7471B

INSTRUMENT ID: CV1

 μ Abs.:

54893



Linear

A= 0.0000e+000

B= 1.8187e-004

C= -5.3528e-002

Rho= 0.9997806

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.040	-0.040	72	0.000	72				
0.2	0.200	0.160	-0.040	1174	0.0 %	1174				
2.5	2.500	2.509	0.009	14089	0.0 %	14089				
5.0	5.000	5.165	0.165	28692	0.0 %	28692				
7.5	7.500	7.478	-0.022	41410	0.0 %	41410				
10.0	10.000	9.930	-0.070	54893	0.0 %	54893				

LB100332 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Jan 2019 13:47:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	72	-	0.0000	7471B	PPB	16 Jan 2019 14:00:49	S	Std
0.2 - 1	50.2	1174	-	0.2000	7471B	PPB	16 Jan 2019 14:03:08	S	Std
2.5 - 1	52.5	14089	-	2.5000	7471B	PPB	16 Jan 2019 14:05:26	S	Std
5.0 - 1	55.0	28692	-	5.0000	7471B	PPB	16 Jan 2019 14:07:45	S	Std
7.5 - 1	57.5	41410	-	7.5000	7471B	PPB	16 Jan 2019 14:10:02	S	Std
10.0 - 1	510.0	54893	-	10.0000	7471B	PPB	16 Jan 2019 14:12:20	S	Std
ICV65 - 1	ICV65	22827	4.0979	-7471B	PPB	PPB	16 Jan 2019 14:15:31	U	SMPL
ICB65 - 1	ICB65	-53	-0.0632	-7471B	PPB	PPB	16 Jan 2019 14:17:47	U	SMPL
CCV51 - 1	CCV51	28454	5.1213	-7471B	PPB	PPB	16 Jan 2019 14:20:03	U	SMPL
CCB51 - 1	CCB51	-53	-0.0632	-7471B	PPB	PPB	16 Jan 2019 14:22:20	U	SMPL
CRA - 1	CRA	1162	0.1578	-7471B	PPB	PPB	16 Jan 2019 14:24:38	U	SMPL
HighStd - 1	HighStd	54572	9.8712	-7471B	PPB	PPB	16 Jan 2019 14:26:53	U	SMPL
ChkStd - 1	ChkStd	38652	6.9759	-7471B	PPB	PPB	16 Jan 2019 14:29:09	U	SMPL
PB116419BL - 1	PBS	-24	-0.0579	-7471B	PPB	PPB	16 Jan 2019 14:31:28	U	SMPL
PB116419BS - 1	LCSS	20114	3.6045	-7471B	PPB	PPB	16 Jan 2019 14:33:47	U	SMPL
K1007-01 - 1	NB-07-011419	719	0.0772	-7471B	PPB	PPB	16 Jan 2019 14:36:04	U	SMPL
K1010-03 - 1	OK-01-011519-A	1131	0.1522	-7471B	PPB	PPB	16 Jan 2019 14:38:21	U	SMPL
K1010-05 - 1	OK-01-011519-B	2862	0.4670	-7471B	PPB	PPB	16 Jan 2019 14:40:39	U	SMPL
K1012-07 - 1	EO-01-011518-A	2725	0.4421	-7471B	PPB	PPB	16 Jan 2019 14:42:58	U	SMPL
K1012-09 - 1	EO-01-011518-B	2515	0.4039	-7471B	PPB	PPB	16 Jan 2019 14:45:14	U	SMPL
CCV52 - 1	CCV52	28537	5.1364	-7471B	PPB	PPB	16 Jan 2019 14:47:30	U	SMPL
CCB52 - 1	CCB52	-61	-0.0646	-7471B	PPB	PPB	16 Jan 2019 14:49:46	U	SMPL
K1012-11 - 1	EO-01-011518-C	2689	0.4355	-7471B	PPB	PPB	16 Jan 2019 14:52:04	U	SMPL
K1147-01 - 1	TP-1	1360	0.1938	-7471B	PPB	PPB	16 Jan 2019 14:54:20	U	SMPL
K1147-04 - 1	TP-2	76100	13.7864	-7471B	PPB	PPB	16 Jan 2019 14:56:36	U	SMPL
K1147-07 - 1	TP-3	175	-0.0217	-7471B	PPB	PPB	16 Jan 2019 14:58:53	U	SMPL
K1147-10 - 1	TP-4	1487	0.2169	-7471B	PPB	PPB	16 Jan 2019 15:01:12	U	SMPL
K1147-13 - 1	TP-5	1884	0.2891	-7471B	PPB	PPB	16 Jan 2019 15:03:30	U	SMPL
K1147-16 - 1	TP-6	1767	0.2678	-7471B	PPB	PPB	16 Jan 2019 15:05:47	U	SMPL
K1147-19 - 1	TP-7	503	0.0380	-7471B	PPB	PPB	16 Jan 2019 15:08:04	U	SMPL
K1147-22 - 1	TP-8	801	0.0921	-7471B	PPB	PPB	16 Jan 2019 15:10:22	U	SMPL
K1147-25 - 1	TP-9	755	0.0838	-7471B	PPB	PPB	16 Jan 2019 15:12:39	U	SMPL
CCV53 - 1	CCV53	28049	5.0476	-7471B	PPB	PPB	16 Jan 2019 15:14:55	U	SMPL
CCB53 - 1	CCB53	-77	-0.0675	-7471B	PPB	PPB	16 Jan 2019 15:17:12	U	SMPL
K1147-25DUP - 1	TP-9DUP	934	0.1163	-7471B	PPB	PPB	16 Jan 2019 15:19:29	U	SMPL
K1147-25MS - 1	TP-9MS	21276	3.8158	-7471B	PPB	PPB	16 Jan 2019 15:21:46	U	SMPL
K1147-25MSD - 1	TP-9MSD	20319	3.6418	-7471B	PPB	PPB	16 Jan 2019 15:24:02	U	SMPL
K1147-25LX5 - 1		191	-0.0188	-7471B	PPB	PPB	16 Jan 2019 15:30:45	U	SMPL
K1147-25A - 1		22371	4.0150	-7471B	PPB	PPB	16 Jan 2019 15:33:02	U	SMPL
K1147-04DLX2 - 1		36266	6.5420	-7471B	PPB	PPB	16 Jan 2019 15:35:19	U	SMPL
CCV54 - 1	CCV54	26857	4.8308	-7471B	PPB	PPB	16 Jan 2019 15:37:36	U	SMPL
CCB54 - 1	CCB54	-104	-0.0724	-7471B	PPB	PPB	16 Jan 2019 15:39:56	U	SMPL