



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100335

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/16/2019	14:15
2	CAL2	1	Water	NA	NA	20.2	7.00	01/16/2019	14:16
3	CAL3	1	Water	NA	NA	20.2	10.02	01/16/2019	14:17
4	ICV	1	Water	NA	NA	20.2	6.99	01/16/2019	14:18
5	CCV1	1	Water	NA	NA	20.2	2.01	01/16/2019	14:19
6	K1148-02	1	Water	NA	NA	12.4	6.70	01/16/2019	14:20
7	K1148-02DUP	1	Water	NA	NA	12.5	6.71	01/16/2019	14:21
8	K1148-05	1	Water	NA	NA	13.8	6.51	01/16/2019	14:22
9	CCV2	1	Water	NA	NA	20.1	12.01	01/16/2019	14:23

LB 100335

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity W K1148

WorkList ID : 121368

Date : 1/16/2019 2:06:46 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/16/2019	Water	K1148-02	Corrosivity	Cool 4 deg C	PSEG01	K21	T-1	01/16/2019	9040C
01/16/2019	Water	K1148-05	Corrosivity	Cool 4 deg C	PSEG01	K21	T-2	01/16/2019	9040C

Date/Time 01-16-19 2:10 PM
Received by: JS
Relinquished by: LS

Date/Time 01-16-19 3:00 PM
Received by: CS
Relinquished by: JS