



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: AparnaDate: 01/17/2019OVEN TEMP1: 103 °C TEMP: 105 °CDATE & TIME IN: 01/16/2019 16:48 OUT: 01/17/2019 08:00OVEN TEMP2: 105 °C TEMP: 105 °CDATE & TIME IN: 01/17/2019 09:00 OUT: 01/17/2019 10:35Run Number: LB100343BalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB100343BL	LB100343BL	82.5698	82.5698	100	82.5698	82.5698	82.5698	0.0000	0
2	K1166-01	D-10	74.7931	74.7931	100	74.8440	74.8440	74.8440	0.0509	509
3	K1166-01DUP	D-10DUP	84.5586	84.5586	100	84.6083	84.6083	84.6083	0.0497	497
4	K1166-02	CLOVER-YARD	104.6932	104.6932	100	104.7150	104.7150	104.7150	0.0218	218

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ts011619

WorkList ID : 121374

Date : 1/16/2019 4:40:43 PM

23106343

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/23/2019	Water	K1166-01	✓ TS	Cool 4 deg C	TULL02	D-10	D-10	01/16/2019	SM2540 B
01/23/2019	Water	K1166-02	✓ TS	Cool 4 deg C	TULL02	CLOVER-YARD	CLOVER-YARD	01/16/2019	SM2540 B

Date/Time

Received by:

Relinquished by:

01/17/19, 16:20
AS
CS

Date/Time

Received by:

Relinquished by:

01/17/19, 17:00
AS
AS