

LB100352


INSTRUMENT ID: CV1

245.1

Linear

 μ Abs.:

35841



A= 0.0000e+000

B= 2.8062e-004

C= -3.6095e-002

Rho= 0.9999537

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	92	0.000	92	0			
0.05	0.050	0.052	0.002	314	0.0 %	314				
0.2	0.200	0.181	-0.019	773	0.0 %	773				
2.5	2.500	2.530	0.030	9143	0.0 %	9143				
5.0	5.000	5.047	0.047	18113	0.0 %	18113				
7.5	7.500	7.430	-0.070	26605	0.0 %	26605				
10.0	10.000	10.021	0.021	35841	0.0 %	35841				

LB100352 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 17 Jan 2019 10:18:37

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	92	-	0.0000	245.1	PPB	17 Jan 2019 10:21:54	S	Std
0.05 - 1	50-05	314	-	0.0500	245.1	PPB	17 Jan 2019 10:24:12	S	Std
0.2 - 1	50-2	773	-	0.2000	245.1	PPB	17 Jan 2019 10:26:30	S	Std
2.5 - 1	52-5	9143	-	2.5000	245.1	PPB	17 Jan 2019 10:28:48	S	Std
5.0 - 1	55-0	18113	-	5.0000	245.1	PPB	17 Jan 2019 10:31:05	S	Std
7.5 - 1	57-5	26605	-	7.5000	245.1	PPB	17 Jan 2019 10:33:22	S	Std
10.0 - 1	510-0	35841	-	10.0000	245.1	PPB	17 Jan 2019 11:08:05	S	Std
ICV66 - 1	ICV66	14837	4.1274	-245.1		PPB	17 Jan 2019 11:22:13	U	SMPL
ICB66 - 1	ICB66	-12	-0.0395	-245.1		PPB	17 Jan 2019 11:24:28	U	SMPL
CCV55 - 1	CCV55	18046	5.0279	-245.1		PPB	17 Jan 2019 11:26:45	U	SMPL
CCB55 - 1	CCB55	-35	-0.0459	-245.1		PPB	17 Jan 2019 11:29:01	U	SMPL
CRA - 1	CRA	290	0.0453	-245.1		PPB	17 Jan 2019 11:31:18	U	SMPL
HighStd - 1	HighStd	36478	10.2002	-245.1		PPB	17 Jan 2019 11:33:34	U	SMPL
ChkStd - 1	ChkStd	24870	6.9428	-245.1		PPB	17 Jan 2019 11:35:50	U	SMPL
PB116432BL - 1	PBW	-184	-0.0877	-245.1		PPB	17 Jan 2019 11:38:09	U	SMPL
PB116432BS - 1	LCSW	14920	4.1507	-245.1		PPB	17 Jan 2019 11:40:27	U	SMPL
K1143-01 - 1	0-FLUSH	-146	-0.0771	-245.1		PPB	17 Jan 2019 11:42:44	U	SMPL
K1143-01DUP - 1	0-FLUSHDUP	-58	-0.0524	-245.1		PPB	17 Jan 2019 11:45:01	U	SMPL
K1143-01MS - 1	0-FLUSHMS	14212	3.9520	-245.1		PPB	17 Jan 2019 11:47:19	U	SMPL
K1143-01MSD - 1	0-FLUSHMSD	15219	4.2346	-245.1		PPB	17 Jan 2019 11:49:37	U	SMPL
K1143-02 - 1	1-MIN-FLUSH	-92	-0.0619	-245.1		PPB	17 Jan 2019 11:51:52	U	SMPL
CCV56 - 1	CCV56	18317	5.1039	-245.1		PPB	17 Jan 2019 11:54:09	U	SMPL
CCB56 - 1	CCB56	-22	-0.0423	-245.1		PPB	17 Jan 2019 11:56:25	U	SMPL
K1152-01 - 1	011419-DBRI-E-D-M	-22	-0.0423	-245.1		PPB	17 Jan 2019 11:58:42	U	SMPL
K1152-02 - 1	011419-DBRI-E-U-M	-18	-0.0411	-245.1		PPB	17 Jan 2019 12:00:59	U	SMPL
K1143-01LX5 - 1		47	-0.0229	-245.1		PPB	17 Jan 2019 12:03:15	U	SMPL
K1143-01A - 1		15511	4.3165	-245.1		PPB	17 Jan 2019 12:05:32	U	SMPL
CCV57 - 1	CCV57	18619	5.1887	-245.1		PPB	17 Jan 2019 12:07:49	U	SMPL
CCB57 - 1	CCB57	-17	-0.0409	-245.1		PPB	17 Jan 2019 12:10:06	U	SMPL