

LB100354


INSTRUMENT ID: CV1

7470A

Linear

 μ Abs.:

35692



A= 0.0000e+000

B= 2.7702e-004

C= -4.3259e-002

Rho= 0.9995351

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.022	-0.022	75	0.000	75				
0.2	0.200	0.159	-0.041	730	0.0 %	730				
2.5	2.500	2.554	0.054	9374	0.0 %	9374				
5.0	5.000	4.955	-0.045	18042	0.0 %	18042				
7.5	7.500	7.711	0.211	27992	0.0 %	27992				
10.0	10.000	9.844	-0.156	35692	0.0 %	35692	0			

LB100354 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Jan 2019 13:28:12

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	75	-	0.00007470A	PPB		17 Jan 2019 13:29:31	S	Std
0.2 - 1	50.2	730	-	0.20007470A	PPB		17 Jan 2019 13:31:41	S	Std
2.5 - 1	52.5	9374	-	2.50007470A	PPB		17 Jan 2019 13:33:49	S	Std
5.0 - 1	55.0	18042	-	5.00007470A	PPB		17 Jan 2019 13:35:58	S	Std
7.5 - 1	57.5	27992	-	7.50007470A	PPB		17 Jan 2019 13:38:06	S	Std
10.0 - 1	510.0	35692	-	10.00007470A	PPB		17 Jan 2019 13:40:13	S	Std
ICV67 - 1	ICV67	15029	4.1201	-7470A	PPB		17 Jan 2019 13:43:46	U	SMPL
ICB67 - 1	ICB67	31	-0.0347	-7470A	PPB		17 Jan 2019 13:45:52	U	SMPL
CCV58 - 1	CCV58	19876	5.4628	-7470A	PPB		17 Jan 2019 13:47:59	U	SMPL
CCB58 - 1	CCB58	-38	-0.0538	-7470A	PPB		17 Jan 2019 13:50:06	U	SMPL
CRA - 1	CRA	747	0.1637	-7470A	PPB		17 Jan 2019 13:52:13	U	SMPL
HighStd - 1	HighStd	36101	9.9575	-7470A	PPB		17 Jan 2019 13:54:20	U	SMPL
ChkStd - 1	ChkStd	26750	7.3670	-7470A	PPB		17 Jan 2019 13:56:27	U	SMPL
PB116430BL - 1	PBW	60	-0.0266	-7470A	PPB		17 Jan 2019 13:58:35	U	SMPL
PB116430BS - 1	LCSW	12718	3.4799	-7470A	PPB		17 Jan 2019 14:00:44	U	SMPL
K1010-04 - 1	OK-01-011519-A	8	-0.0410	-7470A	PPB		17 Jan 2019 14:02:52	U	SMPL
K1010-04DUP - 1	OK-01-011519-ADUP	74	-0.0228	-7470A	PPB		17 Jan 2019 14:05:00	U	SMPL
K1010-04MS - 1	OK-01-011519-AMS	12254	3.3513	-7470A	PPB		17 Jan 2019 14:07:08	U	SMPL
K1010-04MSD - 1	OK-01-011519-AMSD	12683	3.4702	-7470A	PPB		17 Jan 2019 14:09:17	U	SMPL
K1010-06 - 1	OK-01-011519-B	138	-0.0050	-7470A	PPB		17 Jan 2019 14:11:24	U	SMPL
CCV59 - 1	CCV59	19413	5.3345	-7470A	PPB		17 Jan 2019 14:13:31	U	SMPL
CCB59 - 1	CCB59	4	-0.0422	-7470A	PPB		17 Jan 2019 14:15:38	U	SMPL
K1012-08 - 1	EO-01-011518-A	127	-0.0081	-7470A	PPB		17 Jan 2019 14:17:45	U	SMPL
K1012-10 - 1	EO-01-011518-B	100	-0.0156	-7470A	PPB		17 Jan 2019 14:19:52	U	SMPL
K1012-12 - 1	EO-01-011518-C	73	-0.0230	-7470A	PPB		17 Jan 2019 14:21:59	U	SMPL
K1147-03 - 1	TP-1	28	-0.0355	-7470A	PPB		17 Jan 2019 14:24:06	U	SMPL
K1147-06 - 1	TP-2	169	0.0036	-7470A	PPB		17 Jan 2019 14:26:13	U	SMPL
K1147-09 - 1	TP-3	167	0.0030	-7470A	PPB		17 Jan 2019 14:28:21	U	SMPL
K1147-12 - 1	TP-4	280	0.0343	-7470A	PPB		17 Jan 2019 14:30:29	U	SMPL
K1147-15 - 1	TP-5	147	-0.0025	-7470A	PPB		17 Jan 2019 14:32:38	U	SMPL
K1147-18 - 1	TP-6	183	0.0074	-7470A	PPB		17 Jan 2019 14:34:47	U	SMPL
K1147-21 - 1	TP-7	117	-0.0108	-7470A	PPB		17 Jan 2019 14:36:55	U	SMPL
CCV60 - 1	CCV60	18228	5.0063	-7470A	PPB		17 Jan 2019 14:39:02	U	SMPL
CCB60 - 1	CCB60	22	-0.0372	-7470A	PPB		17 Jan 2019 14:41:09	U	SMPL
K1147-24 - 1	TP-8	62	-0.0261	-7470A	PPB		17 Jan 2019 14:43:17	U	SMPL
K1147-27 - 1	TP-9	72	-0.0233	-7470A	PPB		17 Jan 2019 14:45:24	U	SMPL
PB116431BL - 1	PBW	75	-0.0225	-7470A	PPB		17 Jan 2019 14:47:31	U	SMPL
PB116431BS - 1	LCSW	14477	3.9672	-7470A	PPB		17 Jan 2019 14:49:38	U	SMPL
K1149-01 - 1	SW001-190114-01	-3	-0.0441	-7470A	PPB		17 Jan 2019 14:51:45	U	SMPL
K1149-02 - 1	SW002-190114-01	167	0.0030	-7470A	PPB		17 Jan 2019 14:53:53	U	SMPL
K1149-03 - 1	SW003-190114-01	71	-0.0236	-7470A	PPB		17 Jan 2019 14:56:01	U	SMPL
K1149-04 - 1	SW004-190114-01	-2	-0.0438	-7470A	PPB		17 Jan 2019 14:58:09	U	SMPL
K1149-04DUP - 1	SW004-190114-01DUP	13	-0.0397	-7470A	PPB		17 Jan 2019 15:00:18	U	SMPL
K1149-04MS - 1	SW004-190114-01MS	13624	3.7309	-7470A	PPB		17 Jan 2019 15:02:25	U	SMPL
CCV61 - 1	CCV61	17755	4.8752	-7470A	PPB		17 Jan 2019 15:04:32	U	SMPL
CCB61 - 1	CCB61	-9	-0.0458	-7470A	PPB		17 Jan 2019 15:06:40	U	SMPL
K1149-04MSD - 1	SW004-190114-01MSD	13380	3.6633	-7470A	PPB		17 Jan 2019 15:08:47	U	SMPL
PB116430TB - 1	PB116404TB	69	-0.0241	-7470A	PPB		17 Jan 2019 15:10:54	U	SMPL
K1010-04LX5 - 1		76	-0.0222	-7470A	PPB		17 Jan 2019 15:13:02	U	SMPL
K1010-04A - 1		15146	4.1525	-7470A	PPB		17 Jan 2019 15:15:10	U	SMPL
K1149-04LX5 - 1		-8	-0.0455	-7470A	PPB		17 Jan 2019 15:17:17	U	SMPL
K1149-04A - 1		15311	4.1982	-7470A	PPB		17 Jan 2019 15:19:25	U	SMPL
CCV62 - 1	CCV62	17925	4.9223	-7470A	PPB		17 Jan 2019 15:21:34	U	SMPL
CCB62 - 1	CCB62	-7	-0.0452	-7470A	PPB		17 Jan 2019 15:23:42	U	SMPL