



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : AMANDEEP

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100357

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/17/2019	09:57
2	CAL2	1	Water	NA	NA	20.2	7.00	01/17/2019	09:58
3	CAL3	1	Water	NA	NA	20.3	10.02	01/17/2019	09:59
4	ICV	1	Water	NA	NA	20.2	6.99	01/17/2019	10:00
5	CCV1	1	Water	NA	NA	20.2	2.01	01/17/2019	10:01
6	K1170-01	1	Water	NA	NA	20.5	4.62	01/17/2019	10:03
7	K1170-01DUP	1	Water	NA	NA	20.5	4.60	01/17/2019	10:04
8	CCV2	1	Water	NA	NA	20.1	12.01	01/17/2019	10:05

WORKLIST(Hardcopy Internal Chain)

CB100357

WorkList Name : ph011719

WorkList ID : 121417

Date : 1/17/2019 09:53:44 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/15/2019	Water	K1170-01	pH	Cool 4 deg C	STER14	J61	JS-SYSTEM	01/15/2019	SM 4500-H B

Date/Time 01-17-19 9:55
 Received by: AR
 Relinquished by: CP

Date/Time 01-17-19 10:10
 Received by: CP
 Relinquished by: AR