



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: AMANDEEP

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB100360

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.9	01/17/2019	10:20
2	K1148-02		1	8	>100.0	>212.0	>212.0	01/17/2019	10:26
3	K1148-05		1	7	>100.0	>212.0	>212.0	01/17/2019	10:35
4	K1166-01		1	9	>100.0	>212.0	>212.0	01/17/2019	10:41
5	K1166-02		1	6	>100.0	>212.0	>212.0	01/17/2019	10:46
6	K1166-02DUP		1	8	>100.0	>212.0	>212.0	01/17/2019	10:53

$$\text{Result} = (\text{Celsius} * 1.8) + 32$$

$$\text{Final Result} = \text{Result} + (760 - \text{Ambient Barometric Pressure}) * 0.06$$

WORKLIST(Hardcopy Internal Chain)

CB100360

WorkList Name : FLASH011719

WorkList ID : 121401

Date : 1/17/2019 9:33:20 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/21/2019	Water	K1148-02	Flash Point	Cool 4 deg C	PSEG01	K21	T-1	01/16/2019	1010A
01/21/2019	Water	K1148-05	Flash Point	Cool 4 deg C	PSEG01	K21	T-2	01/16/2019	1010A
01/29/2019	Water	K1166-01	Flash Point	Cool 4 deg C	TULL02		D-10	01/16/2019	1010A
01/29/2019	Water	K1166-02	Flash Point	Cool 4 deg C	TULL02		CLOVER-YARD	01/16/2019	1010A

Date/Time 01-17-19 9:35
 Received by: AR
 Relinquished by: CR

Date/Time 01-17-19 11:00
 Received by: CR
 Relinquished by: AR