



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Aparna

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100366

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/17/2019	09:57
2	CAL2	1	Water	NA	NA	20.2	7.00	01/17/2019	09:58
3	CAL3	1	Water	NA	NA	20.2	10.02	01/17/2019	09:59
4	ICV	1	Water	NA	NA	20.2	6.99	01/17/2019	10:00
5	CCV1	1	Water	NA	NA	20.3	2.01	01/17/2019	13:12
6	K1186-01	1	Water	NA	NA	18.7	4.26	01/17/2019	13:15
7	K1186-01DUP	1	Water	NA	NA	18.8	4.25	01/17/2019	13:18
8	K1186-02	1	Water	NA	NA	19.0	4.51	01/17/2019	13:21
9	CCV2	1	Water	NA	NA	20.1	12.01	01/17/2019	13:24