

LB100368


7471B

INSTRUMENT ID : CV1

Linear

 μ Abs.:

56888



A= 0.0000e+000

B= 1.7682e-004

C= -4.6580e-002

Rho= 0.9998931

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.022	0.022	386	0.000	386				
0.2	0.200	0.167	-0.033	1206	0.0 %	1206				
2.5	2.500	2.477	-0.023	14273	0.0 %	14273				
5.0	5.000	5.097	0.097	29091	0.0 %	29091				
7.5	7.500	7.424	-0.076	42251	0.0 %	42251				
10.0	10.000	10.013	0.013	56888	0.0 %	56888				

LB100368 INSTRUMENT ID: CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 Jan 2019 15:54:53

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	386	-	0.0000	7471B	PPB	17 Jan 2019 15:55:16	S	Std
0.2 - 1	50-2	1206	-	0.2000	7471B	PPB	17 Jan 2019 15:57:34	S	Std
2.5 - 1	52-5	14273	-	2.5000	7471B	PPB	17 Jan 2019 15:59:52	S	Std
5.0 - 1	55-0	29091	-	5.0000	7471B	PPB	17 Jan 2019 16:02:10	S	Std
7.5 - 1	57-5	42251	-	7.5000	7471B	PPB	17 Jan 2019 16:04:27	S	Std
10.0 - 1	510-0	56888	-	10.0000	7471B	PPB	17 Jan 2019 16:10:51	S	Std
ICV68 - 1	ICV68	23104	4.0388	-	7471B	PPB	17 Jan 2019 16:14:01	U	SMPL
ICB68 - 1	ICB68	224	-0.0070	-	7471B	PPB	17 Jan 2019 16:16:17	U	SMPL
CCV63 - 1	CCV63	28505	4.9938	-	7471B	PPB	17 Jan 2019 16:18:34	U	SMPL
CCB63 - 1	CCB63	264	0.0001	-	7471B	PPB	17 Jan 2019 16:20:50	U	SMPL
CRA - 1	CRA	1173	0.1608	-	7471B	PPB	17 Jan 2019 16:23:07	U	SMPL
HighStd - 1	HighStd	56323	9.9127	-	7471B	PPB	17 Jan 2019 16:25:23	U	SMPL
ChkStd - 1	ChkStd	40343	7.0870	-	7471B	PPB	17 Jan 2019 16:27:39	U	SMPL
PB116449BL - 1	PBS	211	-0.0093	-	7471B	PPB	17 Jan 2019 16:29:58	U	SMPL
PB116449BS - 1	LCSS	20468	3.5727	-	7471B	PPB	17 Jan 2019 16:32:16	U	SMPL
K1153-01 - 1	RBR200044-RBR200027	1384	0.1981	-	7471B	PPB	17 Jan 2019 16:34:33	U	SMPL
K1153-01DUP - 1	RBR200044-RBR200027DUP	1455	0.2107	-	7471B	PPB	17 Jan 2019 16:36:51	U	SMPL
K1153-01MS - 1	RBR200044-RBR200027MS	21840	3.8153	-	7471B	PPB	17 Jan 2019 16:42:16	U	SMPL
K1153-01MSD - 1	RBR200044-RBR200027MSD	21915	3.8285	-	7471B	PPB	17 Jan 2019 16:44:33	U	SMPL
K1156-08 - 1	COMP-5	17235	3.0010	-	7471B	PPB	17 Jan 2019 16:46:50	U	SMPL
CCV64 - 1	CCV64	29175	5.1123	-	7471B	PPB	17 Jan 2019 16:49:07	U	SMPL
CCB64 - 1	CCB64	248	-0.0027	-	7471B	PPB	17 Jan 2019 16:51:23	U	SMPL
K1156-10 - 1	COMP-7	9178	1.5763	-	7471B	PPB	17 Jan 2019 16:53:40	U	SMPL
K1156-12 - 1	COMP-9	12169	2.1052	-	7471B	PPB	17 Jan 2019 16:55:56	U	SMPL
K1157-01 - 1	WC-01-011519	5540	0.9330	-	7471B	PPB	17 Jan 2019 16:58:12	U	SMPL
K1157-03 - 1	WC-02-011519	24018	4.2004	-	7471B	PPB	17 Jan 2019 17:00:29	U	SMPL
K1157-05 - 1	WC-03-011519	10374	1.7878	-	7471B	PPB	17 Jan 2019 17:02:45	U	SMPL
K1157-07 - 1	WC-04-011519	21962	3.8368	-	7471B	PPB	17 Jan 2019 17:05:04	U	SMPL
K1157-09 - 1	WC-05-011519	5237	0.8794	-	7471B	PPB	17 Jan 2019 17:07:21	U	SMPL
K1157-11 - 1	WC-06-011519	5720	0.9649	-	7471B	PPB	17 Jan 2019 17:09:40	U	SMPL
K1157-13 - 1	WC-07-011519	5351	0.8996	-	7471B	PPB	17 Jan 2019 17:11:58	U	SMPL
K1167-01 - 1	JDHS-SOUTH-1	1197	0.1651	-	7471B	PPB	17 Jan 2019 17:14:14	U	SMPL
CCV65 - 1	CCV65	29280	5.1308	-	7471B	PPB	17 Jan 2019 17:16:30	U	SMPL
CCB65 - 1	CCB65	252	-0.0020	-	7471B	PPB	17 Jan 2019 17:18:47	U	SMPL
K1167-03 - 1	JDHS-NORTH-2	52132	9.1716	-	7471B	PPB	17 Jan 2019 17:21:04	U	SMPL
K1167-05 - 1	D-9	52766	9.2837	-	7471B	PPB	17 Jan 2019 17:23:20	U	SMPL
K1167-07 - 1	D-10	9088	1.5604	-	7471B	PPB	17 Jan 2019 17:25:39	U	SMPL
K1167-09 - 1	CLOVER-YARD-A	26150	4.5774	-	7471B	PPB	17 Jan 2019 17:27:58	U	SMPL
K1167-11 - 1	CLOVER-YARD-B	23575	4.1221	-	7471B	PPB	17 Jan 2019 17:30:15	U	SMPL
K1153-01LX5 - 1		447	0.0325	-	7471B	PPB	17 Jan 2019 17:32:32	U	SMPL
K1153-01A - 1		24742	4.3284	-	7471B	PPB	17 Jan 2019 17:34:50	U	SMPL
CCV66 - 1	CCV66	29415	5.1547	-	7471B	PPB	17 Jan 2019 17:37:08	U	SMPL
CCB66 - 1	CCB66	260	-0.0006	-	7471B	PPB	17 Jan 2019 17:39:26	U	SMPL