

2B100370

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

1/17/2019 16:20

Test: Phenolics- *mg/L*

Sample Id	Result	Dil. 1 +	Response	Errors

ICV1	1.000	0.0	0.116	
ICB1	-0.001	0.0	0.011	
CCV1	1.010	0.0	0.117	
CCB1	-0.000	0.0	0.011	
PB116468BL	-0.002	0.0	0.011	
PB116468BS	1.005	0.0	0.117	
K1166-01	0.018	0.0	0.013	
K1166-01DUP	0.017	0.0	0.013	
K1166-01MS	0.967	0.0	0.113	
K1166-01MSD	0.965	0.0	0.113	
K1166-02	0.029	0.0	0.014	
CCV2	1.011	0.0	0.117	
CCB2	-0.001	0.0	0.011	
N	13			
Mean	0.463			
SD	0.5110			
CV%	110.37			

Aquakem v. 7.2AQ1

Results from time period:

Thu Jan 17 15:48:29 2019

Thu Jan 17 16:15:20 2019

AS

Sample Id	Sam/Ctr/c/	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0492	mg/l	1/17/2019 15:48:29	
0.1PPM	A	Phenolics-	P	0.1002	mg/l	1/17/2019 15:48:30	
0.25PPM	A	Phenolics-	P	0.2492	mg/l	1/17/2019 15:48:31	
0.50PPM	A	Phenolics-	P	0.4989	mg/l	1/17/2019 15:48:32	
1.0PPM	A	Phenolics-	P	1.0042	mg/l	1/17/2019 15:48:33	
2.0PPM	A	Phenolics-	P	1.9983	mg/l	1/17/2019 15:48:34	
ICV1	S	Phenolics-	P	1.0003	mg/l	1/17/2019 16:12:32	
ICB1	S	Phenolics-	P	-0.001	mg/l	1/17/2019 16:12:33	
CCV1	S	Phenolics-	P	1.0099	mg/l	1/17/2019 16:12:34	
CCB1	S	Phenolics-	P	-0.0001	mg/l	1/17/2019 16:12:35	
PB116468BL	S	Phenolics-	P	-0.002	mg/l	1/17/2019 16:12:36	
PB116468BS	S	Phenolics-	P	1.0047	mg/l	1/17/2019 16:12:37	
K1166-01	S	Phenolics-	P	0.0179	mg/l	1/17/2019 16:12:38	
K1166-01DUP	S	Phenolics-	P	0.0175	mg/l	1/17/2019 16:12:39	
K1166-01MS	S	Phenolics-	P	0.9674	mg/l	1/17/2019 16:12:40	
K1166-01MSD	S	Phenolics-	P	0.9646	mg/l	1/17/2019 16:12:41	
K1166-02	S	Phenolics-	P	0.0294	mg/l	1/17/2019 16:12:42	
CCV2	S	Phenolics-	P	1.0113	mg/l	1/17/2019 16:12:43	
CCB2	S	Phenolics-	P	-0.0007	mg/l	1/17/2019 16:15:20	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : *CH* 1/17/2019 15:50


Test Phenolics-

Accepted 1/17/2019 15:50

Factor 9.497
Bias 0.011

Coeff. of det. 0.999992

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.016	0.0492	0.0500	
2	Phenol-2	0.022	0.1002	0.1000	
3	Phenol-2	0.037	0.2492	0.2500	
4	Phenol-2	0.063	0.4989	0.5000	
5	Phenol-2	0.117	1.0042	1.0000	
6	Phenol-2	0.221	1.9983	2.0000	