


LB100383
INSTRUMENT ID: CV1

7470A

Linear

μ Abs.:

41039



A= 0.0000e+000

B= 2.4392e-004

C= 6.9040e-003

Rho= 0.9997406

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.027	0.027	82	0.000	82				
0.2	0.200	0.200	0.000	793	0.0 %	793				
2.5	2.500	2.401	-0.099	9816	0.0 %	9816				
5.0	5.000	5.150	0.150	21085	0.0 %	21085				
7.5	7.500	7.404	-0.096	30328	0.0 %	30328				
10.0	10.000	10.017	0.017	41039	0.0 %	41039	0			

LB100383 INSTRUMENT ID: CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Jan 2019 10:44:10

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	82	-	0.0000	7470A	PPB	18 Jan 2019 10:46:46	S	Std
0.2 - 1	50-2	793	-	0.2000	7470A	PPB	18 Jan 2019 10:48:54	S	Std
2.5 - 1	52-5	9816	-	2.5000	7470A	PPB	18 Jan 2019 10:51:03	S	Std
5.0 - 1	55-0	21085	-	5.0000	7470A	PPB	18 Jan 2019 10:53:12	S	Std
7.5 - 1	57-5	30328	-	7.5000	7470A	PPB	18 Jan 2019 10:55:20	S	Std
10.0 - 1	510-0	41039	-	10.0000	7470A	PPB	18 Jan 2019 10:57:27	S	Std
ICV69 - 1	ICV69	15517	3.7918	-	7470A	PPB	18 Jan 2019 11:01:03	U	SMPL
ICB69 - 1	ICB69	37	0.0159	-	7470A	PPB	18 Jan 2019 11:03:09	U	SMPL
CCV67 - 1	CCV67	19844	4.8472	-	7470A	PPB	18 Jan 2019 11:05:16	U	SMPL
CCB67 - 1	CCB67	20	0.0118	-	7470A	PPB	18 Jan 2019 11:07:23	U	SMPL
CRA - 1	CRA	751	0.1901	-	7470A	PPB	18 Jan 2019 11:09:30	U	SMPL
HighStd - 1	HighStd	40714	9.9378	-	7470A	PPB	18 Jan 2019 11:11:36	U	SMPL
ChkStd - 1	ChkStd	27882	6.8079	-	7470A	PPB	18 Jan 2019 11:13:43	U	SMPL
PB116459BL - 1	PBW	-32	-0.0009	-	7470A	PPB	18 Jan 2019 11:15:52	U	SMPL
PB116459BS - 1	LCSW	15828	3.8677	-	7470A	PPB	18 Jan 2019 11:22:47	U	SMPL
K1153-02 - 1	RBR200044-RBR200027	-9	0.0047	-	7470A	PPB	18 Jan 2019 11:24:55	U	SMPL
K1153-02DUP - 1	RBR200044-RBR200027DUP	49	0.0189	-	7470A	PPB	18 Jan 2019 11:27:03	U	SMPL
K1153-02MS - 1	RBR200044-RBR200027MS	12986	3.1744	-	7470A	PPB	18 Jan 2019 11:29:12	U	SMPL
K1153-02MSD - 1	RBR200044-RBR200027MSD	13102	3.2027	-	7470A	PPB	18 Jan 2019 11:31:20	U	SMPL
K1156-14 - 1	COMP-5	22	0.0123	-	7470A	PPB	18 Jan 2019 11:33:27	U	SMPL
CCV68 - 1	CCV68	20097	4.9090	-	7470A	PPB	18 Jan 2019 11:35:34	U	SMPL
CCB68 - 1	CCB68	24	0.0128	-	7470A	PPB	18 Jan 2019 11:37:41	U	SMPL
K1156-16 - 1	COMP-7	100	0.0313	-	7470A	PPB	18 Jan 2019 11:39:48	U	SMPL
K1156-18 - 1	COMP-9	114	0.0347	-	7470A	PPB	18 Jan 2019 11:41:54	U	SMPL
K1157-15 - 1	WC-01-011519	164	0.0469	-	7470A	PPB	18 Jan 2019 11:44:02	U	SMPL
K1157-17 - 1	WC-02-011519	92	0.0293	-	7470A	PPB	18 Jan 2019 11:46:09	U	SMPL
K1157-19 - 1	WC-03-011519	166	0.0474	-	7470A	PPB	18 Jan 2019 11:48:16	U	SMPL
K1157-21 - 1	WC-04-011519	94	0.0298	-	7470A	PPB	18 Jan 2019 11:50:24	U	SMPL
K1157-23 - 1	WC-05-011519	69	0.0237	-	7470A	PPB	18 Jan 2019 11:52:32	U	SMPL
K1157-25 - 1	WC-06-011519	125	0.0374	-	7470A	PPB	18 Jan 2019 11:54:40	U	SMPL
K1157-27 - 1	WC-07-011519	107	0.0330	-	7470A	PPB	18 Jan 2019 11:56:49	U	SMPL
K1159-01 - 1	SSSI-0115-S-TCLP-7	117	0.0354	-	7470A	PPB	18 Jan 2019 11:58:56	U	SMPL
CCV69 - 1	CCV69	20017	4.8894	-	7470A	PPB	18 Jan 2019 12:01:03	U	SMPL
CCB69 - 1	CCB69	18	0.0113	-	7470A	PPB	18 Jan 2019 12:03:10	U	SMPL
K1160-01 - 1	OILY-STONES	92	0.0293	-	7470A	PPB	18 Jan 2019 12:05:17	U	SMPL
K1167-02 - 1	JDHS-SOUTH-1	91	0.0291	-	7470A	PPB	18 Jan 2019 12:07:24	U	SMPL
K1167-04 - 1	JDHS-NORTH-2	137	0.0403	-	7470A	PPB	18 Jan 2019 12:09:31	U	SMPL
K1167-06 - 1	D-9	105	0.0325	-	7470A	PPB	18 Jan 2019 12:11:38	U	SMPL
K1167-08 - 1	D-10	101	0.0315	-	7470A	PPB	18 Jan 2019 12:13:46	U	SMPL
K1167-10 - 1	CLOVER-YARD-A	114	0.0347	-	7470A	PPB	18 Jan 2019 12:15:54	U	SMPL
K1167-12 - 1	CLOVER-YARD-B	121	0.0364	-	7470A	PPB	18 Jan 2019 12:18:02	U	SMPL
PB116460BL - 1	PBW	79	0.0262	-	7470A	PPB	18 Jan 2019 12:20:11	U	SMPL
PB116460BS - 1	LCSW	13622	3.3296	-	7470A	PPB	18 Jan 2019 12:22:20	U	SMPL
K1148-02 - 1	T-1	54	0.0201	-	7470A	PPB	18 Jan 2019 12:24:27	U	SMPL
CCV70 - 1	CCV70	20003	4.8860	-	7470A	PPB	18 Jan 2019 12:26:34	U	SMPL
CCB70 - 1	CCB70	49	0.0189	-	7470A	PPB	18 Jan 2019 12:28:41	U	SMPL
K1148-02DUP - 1	T-1DUP	92	0.0293	-	7470A	PPB	18 Jan 2019 12:30:49	U	SMPL
K1148-02MS - 1	T-1MS	15696	3.8355	-	7470A	PPB	18 Jan 2019 12:35:31	U	SMPL
K1148-02MSD - 1	T-1MSD	15553	3.8006	-	7470A	PPB	18 Jan 2019 12:37:38	U	SMPL
K1148-05 - 1	T-2	14	0.0103	-	7470A	PPB	18 Jan 2019 12:39:45	U	SMPL
PB116491BL - 1	PBW	90	0.0289	-	7470A	PPB	18 Jan 2019 12:41:53	U	SMPL
PB116491BS - 1	LCSW	13870	3.3901	-	7470A	PPB	18 Jan 2019 12:44:01	U	SMPL
K1148-01 - 1	T-1	2	0.0074	-	7470A	PPB	18 Jan 2019 12:46:09	U	SMPL
K1148-01DUP - 1	T-1DUP	43	0.0174	-	7470A	PPB	18 Jan 2019 12:48:17	U	SMPL
K1148-01MS - 1	T-1MS	15538	3.7969	-	7470A	PPB	18 Jan 2019 12:53:39	U	SMPL
K1148-01MSD - 1	T-1MSD	15690	3.8340	-	7470A	PPB	18 Jan 2019 12:55:46	U	SMPL
CCV71 - 1	CCV71	20715	5.0597	-	7470A	PPB	18 Jan 2019 12:57:53	U	SMPL
CCB71 - 1	CCB71	9	0.0091	-	7470A	PPB	18 Jan 2019 13:00:01	U	SMPL
K1148-04 - 1	T-2	73	0.0247	-	7470A	PPB	18 Jan 2019 13:02:08	U	SMPL
PB116459TB - 1	PB116438TB	105	0.0325	-	7470A	PPB	18 Jan 2019 13:04:15	U	SMPL
PB116460TB - 1	PB116440TB	108	0.0332	-	7470A	PPB	18 Jan 2019 13:06:22	U	SMPL
K1153-02LX5 - 1		107	0.0330	-	7470A	PPB	18 Jan 2019 13:08:30	U	SMPL
K1153-02A - 1		16223	3.9640	-	7470A	PPB	18 Jan 2019 13:10:37	U	SMPL
K1148-02LX5 - 1		52	0.0196	-	7470A	PPB	18 Jan 2019 13:12:45	U	SMPL
K1148-02A - 1		16307	3.9845	-	7470A	PPB	18 Jan 2019 13:14:53	U	SMPL
K1148-01LX5 - 1		41	0.0169	-	7470A	PPB	18 Jan 2019 13:17:01	U	SMPL
K1148-01A - 1		16531	4.0391	-	7470A	PPB	18 Jan 2019 13:19:10	U	SMPL
CCV72 - 1	CCV72	20755	5.0695	-	7470A	PPB	18 Jan 2019 13:21:17	U	SMPL
CCB72 - 1	CCB72	43	0.0174	-	7470A	PPB	18 Jan 2019 13:23:25	U	SMPL