

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB100387

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP70044
calibration std. phosphate 1 ppm	WP70984
calibration std. phosphate 0.5 ppm	WP70983
calibration std. phosphate 0.3 ppm	WP70982
calibration std. 0 ppm	WP70979
phosphate ICV std.	WP70986
phosphate CCV std.	WP70985
calibration std. phosphate 0.1 ppm	WP70981
calibration std. phosphate 0.05 ppm	WP70980
Combined reagent	WP70989
11 N sulfuric acid	WP66905
Phenolphthalein indicator	WP66907
Sodium hydroxide, 1N	WP66924

Intercept: -0.0011Slope: 0.6504Regression: 0.999801

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	01/17/2019	16:01
2	CAL2	0.05	1	50	50	0.033	01/17/2019	16:01
3	CAL3	0.10	1	50	50	0.065	01/17/2019	16:02
4	CAL4	0.30	1	50	50	0.185	01/17/2019	16:02
5	CAL5	0.50	1	50	50	0.330	01/17/2019	16:03
6	CAL6	1.00	1	50	50	0.649	01/17/2019	16:03



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB100387

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.325	0.501	01/17/2019	16:04
2	ICB		1	50	50	0.000	0.002	01/17/2019	16:04
3	CCV1	0.50	1	50	50	0.327	0.504	01/17/2019	16:05
4	CCB1		1	50	50	0.000	0.002	01/17/2019	16:05
5	PB116499BL		1	50	50	0.000	0.002	01/17/2019	16:06
6	PB116499BS	0.50	1	50	50	0.349	0.538	01/17/2019	16:07
7	K1179-01		1	50	50	0.977	1.504	01/17/2019	16:07
8	K1179-01DUP		1	50	50	0.932	1.435	01/17/2019	16:08
9	K1179-01MS	0.50	1	50	50	1.350	2.077	01/17/2019	16:09
10	K1179-01MSD	0.50	1	50	50	1.354	2.083	01/17/2019	16:09
11	K1179-01		5	50	50	0.310	0.478	01/17/2019	16:16
12	K1179-01DUP		5	50	50	0.308	0.475	01/17/2019	16:16
13	K1179-01MS	0.50	5	50	50	0.327	0.504	01/17/2019	16:17
14	K1179-01MSD	0.50	5	50	50	0.330	0.509	01/10/2019	16:18
15	CCV2	0.50	1	50	50	0.329	0.508	01/17/2019	16:19
16	CCB2		1	50	50	0.000	0.002	01/17/2019	16:19