



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100390

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	01/18/2019	11:20
2	CAL2	1	Water	NA	NA	21.2	7.00	01/18/2019	11:21
3	CAL3	1	Water	NA	NA	21.2	10.03	01/18/2019	11:22
4	ICV	1	Water	NA	NA	21.2	7.01	01/18/2019	11:23
5	CCV1	1	Water	NA	NA	21.1	2.01	01/18/2019	11:24
6	K1193-01	1	Water	NA	NA	16.0	4.88	01/18/2019	11:25
7	K1193-01DUP	1	Water	NA	NA	16.1	4.89	01/18/2019	11:26
8	CCV2	1	Water	NA	NA	21.1	12.01	01/18/2019	11:27

LB 100390

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph k1193

WorkList ID : 121459

Date : 1/18/2019 11:06:41 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/11/2019	Water	K1193-01-1	pH	Cool 4 deg C	STER10	K52	CELL-A-POOL-WATER(DE-IC	01/11/2019	SM 4500-H B

Date/Time 01-18-19 11:10 AM
Received by: JA
Relinquished by: CO

Date/Time 01-18-19 11:40 AM
Received by: CO
Relinquished by: SO