

LB100396


INSTRUMENT ID: CV1

7471B

Linear

 μ Abs.:

5811E



A= 0.0000e+000

B= 1.7221e-004

C= 3.0450e-002

Rho= 0.9999371

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.045	0.045	83	0.000	83				
0.2	0.200	0.208	0.008	1029	0.0 %	1029				
2.5	2.500	2.483	-0.017	14243	0.0 %	14243				
5.0	5.000	4.920	-0.080	28393	0.0 %	28393				
7.5	7.500	7.505	0.005	43404	0.0 %	43404				
10.0	10.000	10.039	0.039	58118	0.0 %	58118				

LB100396 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Jan 2019 14:58:44

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	83	-	0.0000	7471B	PPB	18 Jan 2019 15:01:51	S	Std
0.2 - 1	50-2	1029	-	0.2000	7471B	PPB	18 Jan 2019 15:04:08	S	Std
2.5 - 1	52-5	14243	-	2.5000	7471B	PPB	18 Jan 2019 15:06:26	S	Std
5.0 - 1	55-0	28393	-	5.0000	7471B	PPB	18 Jan 2019 15:08:44	S	Std
7.5 - 1	57-5	43404	-	7.5000	7471B	PPB	18 Jan 2019 15:11:02	S	Std
10.0 - 1	510-0	58118	-	10.0000	7471B	PPB	18 Jan 2019 15:13:19	S	Std
ICV70 - 1	ICV70	22966	3.9855	-	7471B	PPB	18 Jan 2019 15:16:12	U	SMPL
ICB70 - 1	ICB70	-67	0.0189	-	7471B	PPB	18 Jan 2019 15:18:28	U	SMPL
CCV73 - 1	CCV73	28081	4.8663	-	7471B	PPB	18 Jan 2019 15:20:45	U	SMPL
CCB73 - 1	CCB73	-58	0.0205	-	7471B	PPB	18 Jan 2019 15:23:01	U	SMPL
CRA - 1	CRA	1018	0.2058	-	7471B	PPB	18 Jan 2019 15:25:18	U	SMPL
HighStd - 1	HighStd	57844	9.9919	-	7471B	PPB	18 Jan 2019 15:27:34	U	SMPL
ChkStd - 1	ChkStd	41030	7.0963	-	7471B	PPB	18 Jan 2019 15:29:50	U	SMPL
PB116508BL - 1	PBS	-154	0.0039	-	7471B	PPB	18 Jan 2019 15:32:09	U	SMPL
PB116508BS - 1	LCSS	23185	4.0232	-	7471B	PPB	18 Jan 2019 15:40:00	U	SMPL
K1011-05 - 1	OR-03-011719-A	2453	0.4529	-	7471B	PPB	18 Jan 2019 15:42:16	U	SMPL
K1011-05DUP - 1	OR-03-011719-ADUP	2435	0.4498	-	7471B	PPB	18 Jan 2019 15:44:34	U	SMPL
K1011-05MS - 1	OR-03-011719-AMS	21133	3.6698	-	7471B	PPB	18 Jan 2019 15:46:52	U	SMPL
K1011-05MSD - 1	OR-03-011719-AMSD	21054	3.6562	-	7471B	PPB	18 Jan 2019 15:49:10	U	SMPL
K1011-07 - 1	OR-03-011719-B	1518	0.2919	-	7471B	PPB	18 Jan 2019 15:51:27	U	SMPL
CCV74 - 1	CCV74	28047	4.8605	-	7471B	PPB	18 Jan 2019 15:53:42	U	SMPL
CCB74 - 1	CCB74	-55	0.0210	-	7471B	PPB	18 Jan 2019 15:55:58	U	SMPL
K1015-09 - 1	SU-01-011719-A	3230	0.5867	-	7471B	PPB	18 Jan 2019 15:58:15	U	SMPL
K1015-11 - 1	SU-01-011719-B	1291	0.2528	-	7471B	PPB	18 Jan 2019 16:00:31	U	SMPL
K1173-01 - 1	RBR-200052	277	0.0782	-	7471B	PPB	18 Jan 2019 16:02:47	U	SMPL
K1173-03 - 1	RBR-200050	204	0.0656	-	7471B	PPB	18 Jan 2019 16:05:04	U	SMPL
K1173-05 - 1	RT-5124	2226	0.4138	-	7471B	PPB	18 Jan 2019 16:07:20	U	SMPL
K1173-07 - 1	RT-3425	1026	0.2071	-	7471B	PPB	18 Jan 2019 16:09:37	U	SMPL
K1180-01 - 1	OILY-STONES	101	0.0478	-	7471B	PPB	18 Jan 2019 16:11:55	U	SMPL
K1189-01 - 1	RBR200034	1030	0.2078	-	7471B	PPB	18 Jan 2019 16:14:13	U	SMPL
K1189-03 - 1	RT3422	1160	0.2302	-	7471B	PPB	18 Jan 2019 16:16:32	U	SMPL
K1191-04 - 1	SITE4-EAST	1437	0.2779	-	7471B	PPB	18 Jan 2019 16:18:49	U	SMPL
CCV75 - 1	CCV75	28338	4.9106	-	7471B	PPB	18 Jan 2019 16:21:06	U	SMPL
CCB75 - 1	CCB75	-61	0.0199	-	7471B	PPB	18 Jan 2019 16:23:23	U	SMPL
K1191-08 - 1	SITE4-WEST	4730	0.8450	-	7471B	PPB	18 Jan 2019 16:25:41	U	SMPL
K1011-05LX5 - 1		497	0.1160	-	7471B	PPB	18 Jan 2019 16:27:57	U	SMPL
K1011-05A - 1		26475	4.5898	-	7471B	PPB	18 Jan 2019 16:30:14	U	SMPL
CCV76 - 1	CCV76	27968	4.8469	-	7471B	PPB	18 Jan 2019 16:32:30	U	SMPL
CCB76 - 1	CCB76	-74	0.0177	-	7471B	PPB	18 Jan 2019 16:34:47	U	SMPL