



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB100397

Analyst By : jignesh

Supervisor Review By : apatel

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	01/18/2019	15:00
2	CAL2	1	Water	NA	NA	21.1	6.99	01/18/2019	15:01
3	CAL3	1	Water	NA	NA	21.1	10.02	01/18/2019	15:02
4	ICV	1	Water	NA	NA	21.2	7.00	01/18/2019	15:03
5	CCV1	1	Water	NA	NA	21.1	2.02	01/18/2019	15:04
6	K1200-02	1	Water	NA	NA	11.2	6.68	01/18/2019	15:05
7	K1200-02DUP	1	Water	NA	NA	11.3	6.69	01/18/2019	15:06
8	CCV2	1	Water	NA	NA	21.1	12.01	01/18/2019	15:07

1-B 100397

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivityW K1200 WorkList ID : 121471 Date : 1/18/2019 2:31:28 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/18/2019	Water	K1200-02	Corrosivity	Cool 4 deg C	PSEG01	K31	WS-1	01/18/2019	9040C

Date/Time 01-18-19 2:35 PM
Received by: JP
Relinquished by: CO

Date/Time 01-18-19 3:38 PM
Received by: CO
Relinquished by: JP