



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100399

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	01/18/2019	15:00
2	CAL2	1	Water	NA	NA	21.1	6.99	01/18/2019	15:01
3	CAL3	1	Water	NA	NA	21.1	10.02	01/18/2019	15:02
4	ICV	1	Water	NA	NA	21.2	7.00	01/18/2019	15:03
5	CCV1	1	Water	NA	NA	21.2	2.02	01/18/2019	16:30
6	K1009-02	1	Solid	20.02	20	19.7	8.40	01/18/2019	16:31
7	K1009-02DUP	1	Solid	20.03	20	19.8	8.41	01/18/2019	16:32
8	K1009-04	1	Solid	20.02	20	19.9	8.36	01/18/2019	16:33
9	K1009-06	1	Solid	20.04	20	19.6	8.40	01/18/2019	16:34
10	K1009-08	1	Solid	20.02	20	19.5	8.33	01/18/2019	16:35
11	K1013-02	1	Solid	20.03	20	20.2	6.94	01/18/2019	16:36
12	K1013-04	1	Solid	20.02	20	20.3	7.20	01/18/2019	16:38
13	K1013-06	1	Solid	20.05	20	20.5	8.59	01/18/2019	16:39
14	K1013-08	1	Solid	20.03	20	20.1	7.70	01/18/2019	16:40
15	K1015-10	1	Solid	20.02	20	20.5	7.64	01/18/2019	16:41
16	CCV2	1	Water	NA	NA	21.1	12.01	01/18/2019	16:42
17	K1015-12	1	Solid	20.03	20	20.4	8.40	01/18/2019	16:43
18	CCV3	1	Water	NA	NA	21.1	2.01	01/18/2019	16:45

VB 100399

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosion k1016

WorkList ID : 121470

Date : 1/18/2019 1:09:41 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/18/2019	Solid	K1009-02	Corrosivity	Cool 4 deg C	PSEG05	K21	CL-01-011819-A	01/18/2019	9045D
01/18/2019	Solid	K1009-04	Corrosivity	Cool 4 deg C	PSEG05	K21	CL-01-011819-B	01/18/2019	9045D
01/18/2019	Solid	K1009-06	Corrosivity	Cool 4 deg C	PSEG05	K21	CL-01-011819-C	01/18/2019	9045D
01/18/2019	Solid	K1009-08	Corrosivity	Cool 4 deg C	PSEG05	K21	CL-01-011819-D	01/18/2019	9045D
01/18/2019	Solid	K1013-02	Corrosivity	Cool 4 deg C	PSEG05	K31	PL-01-011819-A	01/18/2019	9045D
01/18/2019	Solid	K1013-04	Corrosivity	Cool 4 deg C	PSEG05	K31	PL-01-011819-B	01/18/2019	9045D
01/18/2019	Solid	K1013-06	Corrosivity	Cool 4 deg C	PSEG05	K31	PL-01-011819-C	01/18/2019	9045D
01/18/2019	Solid	K1013-08	Corrosivity	Cool 4 deg C	PSEG05	K31	PL-01-011819-D	01/18/2019	9045D
01/17/2019	Solid	K1015-10	Corrosivity	Cool 4 deg C	PSEG05		SU-01-011719-A	01/17/2019	9045D
01/17/2019	Solid	K1015-12	Corrosivity	Cool 4 deg C	PSEG05		SU-01-011719-B	01/17/2019	9045D

Date/Time 01-18-19 2:35pm
 Received by: JP
 Relinquished by:

Date/Time 01-18-19 2:30pm
 Received by: G
 Relinquished by: JP