


LB100407

INSTRUMENT ID: CV1

7470A

μ Abs.:

37761



Linear

A= 0.0000e+000

B= 2.6252e-004

C= 3.5984e-002

Rho= 0.9986041

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.038	0.038	6	0.000	6				
0.2	0.200	0.216	0.016	684	0.0 %	684				
2.5	2.500	2.631	0.131	9887	0.0 %	9887				
5.0	5.000	4.619	-0.381	17458	0.0 %	17458				
7.5	7.500	7.747	0.247	29375	0.0 %	29375				
10.0	10.000	9.949	-0.051	37761	0.0 %	37761	0			

LB100407 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Jan 2019 10:56:46

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	6	-	0.0000	7470A	PPB	21 Jan 2019 10:59:36	S	Std
0.2 - 1	50-2	684	-	0.2000	7470A	PPB	21 Jan 2019 11:01:45	S	Std
2.5 - 1	52-5	9887	-	2.5000	7470A	PPB	21 Jan 2019 11:03:53	S	Std
5.0 - 1	55-0	17458	-	5.0000	7470A	PPB	21 Jan 2019 11:06:02	S	Std
7.5 - 1	57-5	29375	-	7.5000	7470A	PPB	21 Jan 2019 11:08:10	S	Std
10.0 - 1	510-0	37761	-	10.0000	7470A	PPB	21 Jan 2019 11:18:30	S	Std
ICV71 - 1	ICV71	15484	4.1008	-	7470A	PPB	21 Jan 2019 11:21:12	U	SMPL
ICB71 - 1	ICB71	-73	0.0168	-	7470A	PPB	21 Jan 2019 11:23:19	U	SMPL
CCV77 - 1	CCV77	18036	4.7708	-	7470A	PPB	21 Jan 2019 11:25:25	U	SMPL
CCB77 - 1	CCB77	-88	0.0129	-	7470A	PPB	21 Jan 2019 11:27:32	U	SMPL
CRA - 1	CRA	662	0.2098	-	7470A	PPB	21 Jan 2019 11:29:38	U	SMPL
HighStd - 1	HighStd	38029	10.0193	-	7470A	PPB	21 Jan 2019 11:31:45	U	SMPL
ChkStd - 1	ChkStd	27791	7.3316	-	7470A	PPB	21 Jan 2019 11:33:52	U	SMPL
PB116496BL - 1	PBW	-113	0.0063	-	7470A	PPB	21 Jan 2019 11:36:01	U	SMPL
PB116496BS - 1	LCSW	15985	4.2323	-	7470A	PPB	21 Jan 2019 11:41:09	U	SMPL
K1011-06 - 1	OR-03-011719-A	-48	0.0234	-	7470A	PPB	21 Jan 2019 11:43:16	U	SMPL
K1011-06DUP - 1	OR-03-011719-ADUP	5	0.0373	-	7470A	PPB	21 Jan 2019 11:45:24	U	SMPL
K1011-06MS - 1	OR-03-011719-AMS	15701	4.1578	-	7470A	PPB	21 Jan 2019 11:50:51	U	SMPL
K1011-06MSD - 1	OR-03-011719-AMSD	15759	4.1730	-	7470A	PPB	21 Jan 2019 11:53:00	U	SMPL
K1011-08 - 1	OR-03-011719-B	-46	0.0239	-	7470A	PPB	21 Jan 2019 11:55:06	U	SMPL
CCV78 - 1	CCV78	18108	4.7897	-	7470A	PPB	21 Jan 2019 11:57:13	U	SMPL
CCB78 - 1	CCB78	-61	0.0200	-	7470A	PPB	21 Jan 2019 11:59:19	U	SMPL
K1173-02 - 1	RBR-200052	-7	0.0341	-	7470A	PPB	21 Jan 2019 12:01:26	U	SMPL
K1173-04 - 1	RBR-200050	13	0.0394	-	7470A	PPB	21 Jan 2019 12:03:33	U	SMPL
K1173-06 - 1	RT-5124	0	0.0360	-	7470A	PPB	21 Jan 2019 12:05:39	U	SMPL
K1173-08 - 1	RT-3425	27	0.0431	-	7470A	PPB	21 Jan 2019 12:07:46	U	SMPL
K1180-02 - 1	OILY-STONES	-8	0.0339	-	7470A	PPB	21 Jan 2019 12:09:54	U	SMPL
K1184-01 - 1	S1	-2	0.0355	-	7470A	PPB	21 Jan 2019 12:12:02	U	SMPL
K1184-02 - 1	S2	0	0.0360	-	7470A	PPB	21 Jan 2019 12:14:10	U	SMPL
K1184-03 - 1	S3	0	0.0360	-	7470A	PPB	21 Jan 2019 12:16:18	U	SMPL
K1184-04 - 1	S4	6	0.0376	-	7470A	PPB	21 Jan 2019 12:18:29	U	SMPL
K1184-05 - 1	S5	8	0.0381	-	7470A	PPB	21 Jan 2019 12:20:36	U	SMPL
CCV79 - 1	CCV79	18034	4.7702	-	7470A	PPB	21 Jan 2019 12:22:43	U	SMPL
CCB79 - 1	CCB79	-57	0.0210	-	7470A	PPB	21 Jan 2019 12:24:51	U	SMPL
K1187-01 - 1	R1	7	0.0378	-	7470A	PPB	21 Jan 2019 12:26:58	U	SMPL
K1189-02 - 1	RBR200034	15	0.0399	-	7470A	PPB	21 Jan 2019 12:29:05	U	SMPL
K1189-04 - 1	RT3422	17	0.0404	-	7470A	PPB	21 Jan 2019 12:31:12	U	SMPL
PB116497BL - 1	PBW	-11	0.0331	-	7470A	PPB	21 Jan 2019 12:33:20	U	SMPL
PB116497BS - 1	LCSW	13475	3.5734	-	7470A	PPB	21 Jan 2019 12:35:27	U	SMPL
K1178-01 - 1	MW-03	33	0.0446	-	7470A	PPB	21 Jan 2019 12:46:18	U	SMPL
K1178-01DUP - 1	MW-03DUP	4	0.0370	-	7470A	PPB	21 Jan 2019 12:48:25	U	SMPL
K1178-01MS - 1	MW-03MS	13102	3.4755	-	7470A	PPB	21 Jan 2019 12:50:34	U	SMPL
K1178-01MSD - 1	MW-03MSD	12342	3.2760	-	7470A	PPB	21 Jan 2019 12:52:42	U	SMPL
K1178-02 - 1	MW-08	-48	0.0234	-	7470A	PPB	21 Jan 2019 12:54:50	U	SMPL
CCV80 - 1	CCV80	17177	4.5453	-	7470A	PPB	21 Jan 2019 13:07:32	U	SMPL
CCB80 - 1	CCB80	-81	0.0147	-	7470A	PPB	21 Jan 2019 13:09:39	U	SMPL
K1178-03 - 1	MW-07	-4	0.0349	-	7470A	PPB	21 Jan 2019 13:11:47	U	SMPL
K1178-04 - 1	DUP	181	0.0835	-	7470A	PPB	21 Jan 2019 13:13:54	U	SMPL
PB116496TB - 1	PB116470TB	-55	0.0215	-	7470A	PPB	21 Jan 2019 13:16:01	U	SMPL
K1011-06LX5 - 1		37	0.0457	-	7470A	PPB	21 Jan 2019 13:18:09	U	SMPL
K1011-06A - 1		13578	3.6005	-	7470A	PPB	21 Jan 2019 13:20:17	U	SMPL
K1178-01LX5 - 1		18	0.0407	-	7470A	PPB	21 Jan 2019 13:29:23	U	SMPL
K1178-01A - 1		13477	3.5739	-	7470A	PPB	21 Jan 2019 13:31:31	U	SMPL
CCV81 - 1	CCV81	19264	5.0931	-	7470A	PPB	21 Jan 2019 13:36:08	U	SMPL
CCB81 - 1	CCB81	-97	0.0105	-	7470A	PPB	21 Jan 2019 13:38:16	U	SMPL