



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 01/21/2019

TEMP1 IN: 103 01/21/2019 07:20 TEMP OUT: 104 01/21/2019 08:30

TEMP2 IN: 104 01/21/2019 09:10 TEMP OUT: 104 01/21/2019 10:20

TEMP3 IN: 105 01/21/2019 11:30 TEMP OUT: 105 01/21/2019 13:00

TEMP4 IN: 105 01/21/2019 13:40 TEMP OUT: 105 01/21/2019 15:20

Run Number: LB100409

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB100409BL	LB100409BL	1.4025	1.4025	100	1.4025	1.4025	1.4025	0.0000	0
2	LB100409BS	LB100409BS	1.4021	1.4021	100	1.4530	1.4530	1.4530	0.0509	509
3	K1192-03	001M-WILLETS-PT-BLVD- (JAN)	1.4012	1.4012	300	1.4073	1.4073	1.4073	0.0061	20.3
4	K1192-03DUP	001M-WILLETS-PT-BLVD- (JAN) D	1.4087	1.4087	300	1.4146	1.4146	1.4146	0.0059	19.7
5	K1192-04	002M-35TH-AVE- (JAN)	1.4023	1.4023	300	1.4211	1.4211	1.4211	0.0188	62.7
6	K1192-05	001M-WILLETS-PT-BLVD- (DEC)	1.3993	1.3993	300	1.4067	1.4067	1.4067	0.0074	24.7
7	K1200-01	WS-1	1.3988	1.3988	50	1.4316	1.4316	1.4316	0.0328	656
8	K1206-01	011819-DBRI-E-D-M	1.4052	1.4052	500	1.4164	1.4164	1.4164	0.0112	22.4
9	K1206-02	011819-DBRI-E-U-M	1.3944	1.3944	500	1.4057	1.4057	1.4057	0.0113	22.6

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS012119

WorkList ID : 121477

Date : 1/21/2019 7:36:21 AM

213100409

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/24/2019	Water	K1192-03 A	TSS	Cool 4 deg C	TULL01	K51	001M-WILLETTS-PT-BLVD-(JA	01/17/2019	SM2540 D
01/24/2019	Water	K1192-04 A	TSS	Cool 4 deg C	TULL01	K51	002M-35TH-AVE-(JAN)	01/17/2019	SM2540 D
01/24/2019	Water	K1192-05 ✓	TSS	Cool 4 deg C	TULL01	K51	001M-WILLETTS-PT-BLVD-(DE	01/17/2019	SM2540 D
01/23/2019	Water	K1200-01 F	TSS	Cool 4 deg C	PSEG01	K31	WS-1	01/18/2019	SM2540 D
01/25/2019	Water	K1206-01 C	TSS	Cool 4 deg C	HDR103	K23	011819-DBRIE-D-M	01/18/2019	SM2540 D
01/25/2019	Water	K1206-02 C	TSS	Cool 4 deg C	HDR103	K23	011819-DBRIE-U-M	01/18/2019	SM2540 D

PP - 14466661

Date/Time 01/21/19, 08:50
 Received by: AS
 Relinquished by: AS

Date/Time 01/21/19, 12:00
 Received by: AS
 Relinquished by: AS