

LB100413


INSTRUMENT ID : CV1

245.1

Linear

 μ Abs.:

35727



A= 0.0000e+000

B= 2.7990e-004

C= 3.4619e-003

Rho= 0.9999890

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.007	0.007	13	0.000	13	0			
0.05	0.050	0.053	0.003	178	0.0 %	178				
0.2	0.200	0.197	-0.003	693	0.0 %	693				
2.5	2.500	2.474	-0.026	8827	0.0 %	8827				
5.0	5.000	5.032	0.032	17967	0.0 %	17967				
7.5	7.500	7.482	-0.018	26719	0.0 %	26719				
10.0	10.000	10.003	0.003	35727	0.0 %	35727				

LB100413 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 21 Jan 2019 14:58:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	13	-	0.0000	245.1	PPB	21 Jan 2019 15:00:16	S	Std
0.05 - 1	50.05	178	-	0.0500	245.1	PPB	21 Jan 2019 15:02:34	S	Std
0.2 - 1	50.2	693	-	0.2000	245.1	PPB	21 Jan 2019 15:04:52	S	Std
2.5 - 1	52.5	8827	-	2.5000	245.1	PPB	21 Jan 2019 15:07:10	S	Std
5.0 - 1	55.0	17967	-	5.0000	245.1	PPB	21 Jan 2019 15:09:27	S	Std
7.5 - 1	57.5	26719	-	7.5000	245.1	PPB	21 Jan 2019 15:11:44	S	Std
10.0 - 1	510.0	35727	-	10.0000	245.1	PPB	21 Jan 2019 15:20:01	S	Std
ICV72 - 1	ICV72	14250	3.9920	-245.1		PPB	21 Jan 2019 15:31:45	U	SMPL
ICB72 - 1	ICB72	-90	-0.0217	-245.1		PPB	21 Jan 2019 15:34:01	U	SMPL
CCV82 - 1	CCV82	17911	5.0168	-245.1		PPB	21 Jan 2019 15:36:18	U	SMPL
CCB82 - 1	CCB82	-102	-0.0251	-245.1		PPB	21 Jan 2019 15:38:34	U	SMPL
CRA - 1	CRA	150	0.0454	-245.1		PPB	21 Jan 2019 15:40:51	U	SMPL
HighStd - 1	HighStd	35475	9.9329	-245.1		PPB	21 Jan 2019 15:43:07	U	SMPL
ChkStd - 1	ChkStd	25874	7.2456	-245.1		PPB	21 Jan 2019 15:45:23	U	SMPL
PB116498BL - 1	PBW	-148	-0.0380	-245.1		PPB	21 Jan 2019 15:47:42	U	SMPL
PB116498BS - 1	LCSW	14188	3.9747	-245.1		PPB	21 Jan 2019 15:55:00	U	SMPL
K1164-01 - 1	011519-BLST-E-D-B	-53	-0.0114	-245.1		PPB	21 Jan 2019 15:57:16	U	SMPL
K1164-02 - 1	011519-BLST-E-D-M	4	0.0046	-245.1		PPB	21 Jan 2019 15:59:34	U	SMPL
K1164-03 - 1	011519-BLST-E-D-S	19	0.0088	-245.1		PPB	21 Jan 2019 16:01:51	U	SMPL
K1164-04 - 1	011519-BLST-E-U-B	27	0.0110	-245.1		PPB	21 Jan 2019 16:04:09	U	SMPL
K1164-05 - 1	011519-BLST-E-U-M	5	0.0049	-245.1		PPB	21 Jan 2019 16:06:25	U	SMPL
CCV83 - 1	CCV83	18666	5.2281	-245.1		PPB	21 Jan 2019 16:08:41	U	SMPL
CCB83 - 1	CCB83	-86	-0.0206	-245.1		PPB	21 Jan 2019 16:10:57	U	SMPL
K1164-06 - 1	011519-BLST-E-U-S	-11	0.0004	-245.1		PPB	21 Jan 2019 16:13:14	U	SMPL
K1164-07 - 1	011519-DBRI-F-D-M	7	0.0054	-245.1		PPB	21 Jan 2019 16:15:30	U	SMPL
K1164-08 - 1	011519-DBRI-F-U-M	3	0.0043	-245.1		PPB	21 Jan 2019 16:17:46	U	SMPL
K1166-01 - 1	D-10	1069	0.3027	-245.1		PPB	21 Jan 2019 16:20:02	U	SMPL
K1166-02 - 1	CLOVER-YARD	585	0.1672	-245.1		PPB	21 Jan 2019 16:22:19	U	SMPL
K1190-01 - 1	011719-DBRI-E-D-M	11	0.0065	-245.1		PPB	21 Jan 2019 16:24:36	U	SMPL
K1190-02 - 1	011719-DBRI-E-U-M	0	0.0035	-245.1		PPB	21 Jan 2019 16:26:53	U	SMPL
K1190-02DUP - 1	011719-DBRI-E-U-MDUP	13	0.0071	-245.1		PPB	21 Jan 2019 16:29:11	U	SMPL
K1190-02MS - 1	011719-DBRI-E-U-MMS	13354	3.7413	-245.1		PPB	21 Jan 2019 16:31:29	U	SMPL
K1190-02MSD - 1	011719-DBRI-E-U-MMSD	12532	3.5112	-245.1		PPB	21 Jan 2019 16:33:45	U	SMPL
CCV84 - 1	CCV84	18676	5.2309	-245.1		PPB	21 Jan 2019 16:36:02	U	SMPL
CCB84 - 1	CCB84	-115	-0.0287	-245.1		PPB	21 Jan 2019 16:38:19	U	SMPL
K1190-02LX5 - 1		3	0.0043	-245.1		PPB	21 Jan 2019 16:40:36	U	SMPL
K1190-02A - 1		15060	4.2188	-245.1		PPB	21 Jan 2019 16:42:52	U	SMPL
CCV85 - 1	CCV85	18867	5.2843	-245.1		PPB	21 Jan 2019 16:45:09	U	SMPL
CCB85 - 1	CCB85	-96	-0.0234	-245.1		PPB	21 Jan 2019 16:47:25	U	SMPL