



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100416

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.2	4.01	01/21/2019	15:30
2	CAL2	1	Water	NA	NA	22.2	7.01	01/21/2019	15:31
3	CAL3	1	Water	NA	NA	22.1	10.04	01/21/2019	15:32
4	ICV	1	Water	NA	NA	22.2	7.00	01/21/2019	15:33
5	CCV1	1	Water	NA	NA	22.1	2.01	01/21/2019	15:35
6	K1212-01	1	Water	NA	NA	10.2	6.50	01/21/2019	15:36
7	K1212-01DUP	1	Water	NA	NA	10.3	6.51	01/21/2019	15:37
8	K1213-01	1	Water	NA	NA	12.6	7.30	01/21/2019	15:38
9	CCV2	1	Water	NA	NA	22.2	12.02	01/21/2019	15:39

WorkList Name : PH K1212

WorkList ID : 121510

Date : 1/21/2019 2:58:51 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/21/2019	Water	K1212-01 - 1A	pH	Cool 4 deg C	PSEG01	K31	FML-124	01/21/2019	9040C
01/21/2019	Water	K1213-01 - 1A	pH	Cool 4 deg C	PSEG01	K21	MANHOLE	01/21/2019	9040C

Date/Time 01-21-19 3:00 PM
Received by: JR
Relinquished by: JR

Date/Time 01-21-19 4:15 PM
Received by: JR
Relinquished by: JR