


LB100420

INSTRUMENT ID: CV1

7471B

Linear



A= 0.0000e+000

B= 1.8663e-004

C= -1.0324e-001

Rho= 0.9990989

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.050	0.050	821	0.000	821				
0.2	0.200	0.252	0.052	1906	0.0 %	1906				
2.5	2.500	2.527	0.027	14093	0.0 %	14093				
5.0	5.000	4.931	-0.069	26973	0.0 %	26973				
7.5	7.500	7.210	-0.290	39187	0.0 %	39187				
10.0	10.000	10.229	0.229	55362	0.0 %	55362				

LB100420 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Jan 2019 08:28:19

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	821	-	0.0000	7471B	PPB	22 Jan 2019 08:47:25	S	Std
0.2 - 1	50-0	1906	-	0.2000	7471B	PPB	22 Jan 2019 08:49:44	S	Std
2.5 - 1	52-0	14093	-	2.5000	7471B	PPB	22 Jan 2019 08:52:02	S	Std
5.0 - 1	55-0	26973	-	5.0000	7471B	PPB	22 Jan 2019 08:54:19	S	Std
7.5 - 1	57-0	39187	-	7.5000	7471B	PPB	22 Jan 2019 08:56:37	S	Std
10.0 - 1	510-0	55362	-	10.0000	7471B	PPB	22 Jan 2019 09:04:09	S	Std
ICV73 - 1	ICV73	21325	3.8767	-	7471B	PPB	22 Jan 2019 09:07:23	U	SMPL
ICB73 - 1	ICB73	654	0.0188	-	7471B	PPB	22 Jan 2019 09:09:38	U	SMPL
CCV86 - 1	CCV86	27399	5.0104	-	7471B	PPB	22 Jan 2019 09:11:56	U	SMPL
CCB86 - 1	CCB86	649	0.0179	-	7471B	PPB	22 Jan 2019 09:14:12	U	SMPL
CRA - 1	CRA	1732	0.2200	-	7471B	PPB	22 Jan 2019 09:16:29	U	SMPL
HighStd - 1	HighStd	55226	10.2039	-	7471B	PPB	22 Jan 2019 09:18:45	U	SMPL
ChkStd - 1	ChkStd	37026	6.8071	-	7471B	PPB	22 Jan 2019 09:21:01	U	SMPL
PB116520BL - 1	PBS	-148	-0.1309	-	7471B	PPB	22 Jan 2019 09:23:21	U	SMPL
PB116520BS - 1	LCSS	21340	3.8795	-	7471B	PPB	22 Jan 2019 09:30:28	U	SMPL
K1009-01 - 1	CL-01-011819-A	2951	0.4475	-	7471B	PPB	22 Jan 2019 09:32:45	U	SMPL
K1009-03 - 1	CL-01-011819-B	3848	0.6149	-	7471B	PPB	22 Jan 2019 09:35:04	U	SMPL
K1009-05 - 1	CL-01-011819-C	8512	1.4854	-	7471B	PPB	22 Jan 2019 09:37:21	U	SMPL
K1009-07 - 1	CL-01-011819-D	4629	0.7607	-	7471B	PPB	22 Jan 2019 09:39:40	U	SMPL
K1013-01 - 1	PL-01-011819-A	997	0.0828	-	7471B	PPB	22 Jan 2019 09:41:56	U	SMPL
CCV87 - 1	CCV87	27333	4.9981	-	7471B	PPB	22 Jan 2019 09:44:12	U	SMPL
CCB87 - 1	CCB87	-125	-0.1266	-	7471B	PPB	22 Jan 2019 09:46:29	U	SMPL
K1013-03 - 1	PL-01-011819-B	3197	0.4934	-	7471B	PPB	22 Jan 2019 09:48:47	U	SMPL
K1013-05 - 1	PL-01-011819-C	879	0.0608	-	7471B	PPB	22 Jan 2019 09:51:03	U	SMPL
K1013-07 - 1	PL-01-011819-D	973	0.0784	-	7471B	PPB	22 Jan 2019 09:53:19	U	SMPL
K1205-01 - 1	C-1	2824	0.4238	-	7471B	PPB	22 Jan 2019 09:55:36	U	SMPL
K1205-02 - 1	C-2	3346	0.5212	-	7471B	PPB	22 Jan 2019 09:57:53	U	SMPL
K1205-03 - 1	C-3	5436	0.9113	-	7471B	PPB	22 Jan 2019 10:00:10	U	SMPL
K1205-04 - 1	C-4	2633	0.3882	-	7471B	PPB	22 Jan 2019 10:02:27	U	SMPL
K1205-05 - 1	C-5	3687	0.5849	-	7471B	PPB	22 Jan 2019 10:04:44	U	SMPL
K1205-06 - 1	C-6	3378	0.5272	-	7471B	PPB	22 Jan 2019 10:07:03	U	SMPL
K1205-06DUP - 1	C-6DUP	3469	0.5442	-	7471B	PPB	22 Jan 2019 10:09:19	U	SMPL
CCV88 - 1	CCV88	26448	4.8329	-	7471B	PPB	22 Jan 2019 10:11:36	U	SMPL
CCB88 - 1	CCB88	639	0.0160	-	7471B	PPB	22 Jan 2019 10:13:52	U	SMPL
K1205-06MS - 1	C-6MS	24517	4.4725	-	7471B	PPB	22 Jan 2019 10:20:44	U	SMPL
K1205-06MSD - 1	C-6MSD	24513	4.4717	-	7471B	PPB	22 Jan 2019 10:23:00	U	SMPL
K1205-06LX5 - 1		1133	0.1082	-	7471B	PPB	22 Jan 2019 10:25:17	U	SMPL
K1205-06A - 1		24916	4.5470	-	7471B	PPB	22 Jan 2019 10:27:35	U	SMPL
CCV89 - 1	CCV89	27087	4.9521	-	7471B	PPB	22 Jan 2019 10:29:52	U	SMPL
CCB89 - 1	CCB89	638	0.0158	-	7471B	PPB	22 Jan 2019 10:32:10	U	SMPL