



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB100432

Ambient Barometric Pressure (mmHg): 776

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	79.6	01/22/2019	09:40
2	K1200-02		1	8	>100.0	>212.0	>212.0	01/22/2019	10:05
3	K1200-02DUP		1	8	>100.0	>212.0	>212.0	01/22/2019	10:35
4	K1201-02		1	7	>100.0	>212.0	>212.0	01/22/2019	11:05
5	K1202-04		1	9	>100.0	>212.0	>212.0	01/22/2019	11:35
6	K1202-06		1	8	>100.0	>212.0	>212.0	01/22/2019	12:10
7	K1204-01		1	7	>100.0	>212.0	>212.0	01/22/2019	12:40
8	K1212-02		1	6	>100.0	>212.0	>212.0	01/22/2019	13:15
9	K1213-02		1	7	>100.0	>212.0	>212.0	01/22/2019	13:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 700432

WorkList Name : FP12219

WorkList ID : 121525

Date : 1/22/2019 9:06:17 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/23/2019	Water	K1200-02 C	Flash Point	Cool 4 deg C	PSEG01	K31	WS-1	01/18/2019	1010A
01/23/2019	Water	K1201-02 C	Flash Point	Cool 4 deg C	PSEG01	K21	COMP-1	01/18/2019	1010A
01/23/2019	Water	K1202-04 E	Flash Point	Cool 4 deg C	PSEG01	K31	BV-COMP-1	01/18/2019	1010A
01/23/2019	Water	K1202-06 E	Flash Point	Cool 4 deg C	PSEG01	K31	23432	01/18/2019	1010A
01/23/2019	Water	K1204-01 E	Flash Point	Cool 4 deg C	PSEG01	K21	TRE-COMP-1	01/18/2019	1010A
01/25/2019	Water	K1212-02 C	Flash Point	Cool 4 deg C	PSEG01	K31	FML-124	01/21/2019	1010A
01/25/2019	Water	K1213-02 C	Flash Point	Cool 4 deg C	PSEG01	K21	MANHOLE	01/21/2019	1010A

Date/Time 01/22/19 9:10 AM
 Received by: KS
 Relinquished by: KS

Date/Time 01/22/19 13:50
 Received by: KS
 Relinquished by: KS