



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100435

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	01/21/2019	10:30
2	CAL2	1	Water	NA	NA	21.2	7.01	01/21/2019	10:31
3	CAL3	1	Water	NA	NA	21.1	10.02	01/21/2019	10:32
4	ICV	1	Water	NA	NA	21.2	6.99	01/21/2019	10:33
5	CCV1	1	Water	NA	NA	21.2	2.01	01/21/2019	10:34
6	K1210-02	1	Solid	20.02	20	19.8	5.72	01/21/2019	10:35
7	K1210-02DUP	1	Solid	20.03	20	19.7	5.73	01/21/2019	10:36
8	CCV2	1	Water	NA	NA	21.1	12.02	01/21/2019	10:37

FB 100435

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity K1210

WorkList ID : 121538

Date : 1/21/2019 10:20:10 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/21/2019	Solid	K1210-02	Corrosivity	Cool 4 deg C	PSEG01	K31	RT-1495	01/21/2019	9045D

Date/Time 01.21.19 10:25 AM
 Received by: Ja
 Relinquished by: CE

Date/Time 01.21.19 11:10 AM
 Received by: CE
 Relinquished by: Ja