


LB100442

7470A

INSTRUMENT ID: CV1

 μ Abs.:

43707



Linear

A= 0.0000e+000

B= 2.2877e-004

C= 2.3065e-002

Rho= 0.9998950

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.057	0.057	148	0.000	148				
0.2	0.200	0.231	0.031	909	0.0 %	909				
2.5	2.500	2.414	-0.086	10453	0.0 %	10453				
5.0	5.000	4.939	-0.061	21487	0.0 %	21487				
7.5	7.500	7.537	0.037	32846	0.0 %	32846				
10.0	10.000	10.022	0.022	43707	0.0 %	43707	0			

LB100442 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Jan 2019 16:17:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	148	-	0.0000	7470A	PPB	22 Jan 2019 16:19:25	S	Std
0.2 - 1	50-2	909	-	0.2000	7470A	PPB	22 Jan 2019 16:21:35	S	Std
2.5 - 1	52-5	10453	-	2.5000	7470A	PPB	22 Jan 2019 16:23:44	S	Std
5.0 - 1	55-0	21487	-	5.0000	7470A	PPB	22 Jan 2019 16:25:53	S	Std
7.5 - 1	57-5	32846	-	7.5000	7470A	PPB	22 Jan 2019 16:28:01	S	Std
10.0 - 1	510-0	43707	-	10.0000	7470A	PPB	22 Jan 2019 16:30:08	S	Std
ICV75 - 1	ICV75	16775	3.8607	-	7470A	PPB	22 Jan 2019 16:33:24	U	SMPL
ICB75 - 1	ICB75	244	0.0789	-	7470A	PPB	22 Jan 2019 16:35:31	U	SMPL
CCV96 - 1	CCV96	22183	5.0978	-	7470A	PPB	22 Jan 2019 16:37:38	U	SMPL
CCB96 - 1	CCB96	67	0.0384	-	7470A	PPB	22 Jan 2019 16:39:45	U	SMPL
CRA - 1	CRA	882	0.2248	-	7470A	PPB	22 Jan 2019 16:41:52	U	SMPL
HighStd - 1	HighStd	42616	9.7723	-	7470A	PPB	22 Jan 2019 16:43:59	U	SMPL
ChkStd - 1	ChkStd	30516	7.0042	-	7470A	PPB	22 Jan 2019 16:46:07	U	SMPL
PB116563BL - 1	PBW	-109	-0.0019	-	7470A	PPB	22 Jan 2019 16:51:07	U	SMPL
PB116563BS - 1	LCSW	17018	3.9163	-	7470A	PPB	22 Jan 2019 16:53:15	U	SMPL
K1212-02 - 1	FML-124	-70	0.0071	-	7470A	PPB	22 Jan 2019 16:55:23	U	SMPL
K1212-02DUP - 1	FML-124DUP	21	0.0279	-	7470A	PPB	22 Jan 2019 16:57:31	U	SMPL
K1212-02MS - 1	FML-124MS	15638	3.6005	-	7470A	PPB	22 Jan 2019 16:59:40	U	SMPL
K1212-02MSD - 1	FML-124MSD	15512	3.5717	-	7470A	PPB	22 Jan 2019 17:01:48	U	SMPL
K1213-02 - 1	MANHOLE	-42	0.0135	-	7470A	PPB	22 Jan 2019 17:03:55	U	SMPL
CCV97 - 1	CCV97	22651	5.2049	-	7470A	PPB	22 Jan 2019 17:06:02	U	SMPL
CCB97 - 1	CCB97	-46	0.0125	-	7470A	PPB	22 Jan 2019 17:08:09	U	SMPL
PB116564BL - 1	PBW	2	0.0235	-	7470A	PPB	22 Jan 2019 17:10:15	U	SMPL
PB116564BS - 1	LCSW	16072	3.6998	-	7470A	PPB	22 Jan 2019 17:12:22	U	SMPL
K1212-01 - 1	FML-124	-60	0.0093	-	7470A	PPB	22 Jan 2019 17:14:29	U	SMPL
K1212-01DUP - 1	FML-124DUP	4	0.0240	-	7470A	PPB	22 Jan 2019 17:16:36	U	SMPL
K1212-01MS - 1	FML-124MS	16876	3.8838	-	7470A	PPB	22 Jan 2019 17:18:44	U	SMPL
K1212-01MSD - 1	FML-124MSD	19890	4.5733	-	7470A	PPB	22 Jan 2019 17:20:52	U	SMPL
K1213-01 - 1	MANHOLE	-55	0.0105	-	7470A	PPB	22 Jan 2019 17:23:00	U	SMPL
CCV98 - 1	CCV98	22097	5.0782	-	7470A	PPB	22 Jan 2019 17:25:08	U	SMPL
CCB98 - 1	CCB98	40	0.0322	-	7470A	PPB	22 Jan 2019 17:32:11	U	SMPL
PB116563TB - 1	PB116545TB	35	0.0311	-	7470A	PPB	22 Jan 2019 17:34:18	U	SMPL
K1212-02LX5 - 1		27	0.0292	-	7470A	PPB	22 Jan 2019 17:36:25	U	SMPL
K1212-02A - 1		16037	3.6918	-	7470A	PPB	22 Jan 2019 17:38:33	U	SMPL
K1212-01LX5 - 1		-36	0.0148	-	7470A	PPB	22 Jan 2019 17:40:40	U	SMPL
K1212-01A - 1		16175	3.7234	-	7470A	PPB	22 Jan 2019 17:42:47	U	SMPL
CCV99 - 1	CCV99	22171	5.0951	-	7470A	PPB	22 Jan 2019 17:44:54	U	SMPL
CCB99 - 1	CCB99	-51	0.0114	-	7470A	PPB	22 Jan 2019 17:47:02	U	SMPL