

LB100446


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

47021



A= 0.0000e+000

B= 2.1755e-004

C= -1.9118e-001

Rho= 0.9999037

Accept=Accepted

slope
+1/015cept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.044	-0.044	675	0.000	675					-
0.2	0.200	0.208	0.008	1834	0.0 %	1834					4
2.5	2.500	2.576	0.076	12721	0.0 %	12721					3
5.0	5.000	5.002	0.002	23872	0.0 %	23872					0
7.5	7.500	7.420	-0.080	34985	0.0 %	34985					0
10.0	10.000	10.038	0.038	47021	0.0 %	47021					-1

nb
1/23/19

LB100446
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	675	-		0 ISM02.4	PPB	1/23/2019 11:15	Std
	0.2 S0.2	1834	-		0.2 ISM02.4	PPB	1/23/2019 11:17	Std
	2.5 S2.5	12721	-		2.5 ISM02.4	PPB	1/23/2019 11:19	Std
	5 S5	23872	-		5 ISM02.4	PPB	1/23/2019 11:22	Std
	7.5 S7.5	34985	-		7.5 ISM02.4	PPB	1/23/2019 11:24	Std
	10 S10.0	47021	-		10 ISM02.4	PPB	1/23/2019 11:26	Std
ICV	ICV	19085	3.9608 -		ISM02.4	PPB	1/23/2019 11:35	SMPL
ICB	ICB	604	-0.0598 -		ISM02.4	PPB	1/23/2019 11:37	SMPL
CCV004	CCV004	23604	4.9439 -		ISM02.4	PPB	1/23/2019 11:39	SMPL
CCB004	CCB004	566	-0.068 -		ISM02.4	PPB	1/23/2019 11:42	SMPL
PB116565BL	PBW565	453	-0.0926 -		ISM02.4	PPB	1/23/2019 11:46	SMPL
K1169-01	METC14	825	-0.0117 -		ISM02.4	PPB	1/23/2019 11:49	SMPL
K1169-02	METC14D	788	-0.0197 -		ISM02.4	PPB	1/23/2019 11:51	SMPL
K1169-03	METC14S	6540	1.2316 -		ISM02.4	PPB	1/23/2019 11:53	SMPL
CCV005	CCV005	23181	4.8519 -		ISM02.4	PPB	1/23/2019 11:56	SMPL
CCB005	CCB005	536	-0.0746 -		ISM02.4	PPB	1/23/2019 11:58	SMPL