

LB100448


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44180



A= 0.0000e+000

B= 2.3021e-004

C= -1.1746e-001

Rho= 0.9998580

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.106	0.106	971	0.000	971				
0.2	0.200	0.165	-0.035	1225	0.0 %	1225				
2.5	2.500	2.425	-0.075	11043	0.0 %	11043				
5.0	5.000	4.998	-0.002	22222	0.0 %	22222				
7.5	7.500	7.453	-0.047	32886	0.0 %	32886				
10.0	10.000	10.053	0.053	44180	0.0 %	44180	0			

LB100448 INSTRUMENT ID: CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Jan 2019 13:05:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	971	-	0.0000	7470A	PPB	23 Jan 2019 13:08:45	S	Std
0.2 - 1	50-2	1225	-	0.2000	7470A	PPB	23 Jan 2019 13:10:54	S	Std
2.5 - 1	52-5	11043	-	2.5000	7470A	PPB	23 Jan 2019 13:13:03	S	Std
5.0 - 1	55-0	22222	-	5.0000	7470A	PPB	23 Jan 2019 13:16:23	S	Std
7.5 - 1	57-5	32886	-	7.5000	7470A	PPB	23 Jan 2019 13:18:31	S	Std
10.0 - 1	510-0	44180	-	10.0000	7470A	PPB	23 Jan 2019 13:27:33	S	Std
ICV77 - 1	ICV77	17631	3.9414	-	7470A	PPB	23 Jan 2019 13:30:36	U	SMPL
ICB77 - 1	ICB77	799	0.0665	-	7470A	PPB	23 Jan 2019 13:32:43	U	SMPL
CCV06 - 1	CCV06	22250	5.0047	-	7470A	PPB	23 Jan 2019 13:34:49	U	SMPL
CCB06 - 1	CCB06	866	0.0819	-	7470A	PPB	23 Jan 2019 13:36:56	U	SMPL
CRA - 1	CRA	1522	0.2329	-	7470A	PPB	23 Jan 2019 13:39:03	U	SMPL
HighStd - 1	HighStd	43361	9.8646	-	7470A	PPB	23 Jan 2019 13:41:10	U	SMPL
ChkStd - 1	ChkStd	32223	7.3006	-	7470A	PPB	23 Jan 2019 13:43:16	U	SMPL
PB116579BL - 1	PBW	611	0.0232	-	7470A	PPB	23 Jan 2019 13:45:25	U	SMPL
PB116579BS - 1	LCSW	18902	4.2340	-	7470A	PPB	23 Jan 2019 13:47:34	U	SMPL
K1210-02 - 1	RT-1495	579	0.0158	-	7470A	PPB	23 Jan 2019 13:49:42	U	SMPL
K1210-02DUP - 1	RT-1495DUP	686	0.0405	-	7470A	PPB	23 Jan 2019 13:51:50	U	SMPL
K1210-02MS - 1	RT-1495MS	15470	3.4439	-	7470A	PPB	23 Jan 2019 13:53:58	U	SMPL
K1210-02MSD - 1	RT-1495MSD	15662	3.4881	-	7470A	PPB	23 Jan 2019 13:56:07	U	SMPL
K1217-01 - 1	DIELECTRIC-FLUIDS	724	0.0492	-	7470A	PPB	23 Jan 2019 13:58:14	U	SMPL
CCV07 - 1	CCV07	22903	5.1550	-	7470A	PPB	23 Jan 2019 14:00:21	U	SMPL
CCB07 - 1	CCB07	868	0.0824	-	7470A	PPB	23 Jan 2019 14:02:28	U	SMPL
PB116580BL - 1	PBW	770	0.0598	-	7470A	PPB	23 Jan 2019 14:04:34	U	SMPL
PB116580BS - 1	LCSW	17405	3.8893	-	7470A	PPB	23 Jan 2019 14:06:42	U	SMPL
K1218-01 - 1	4-3PW-01-1DW	956	0.1026	-	7470A	PPB	23 Jan 2019 14:08:49	U	SMPL
K1218-01DUP - 1	4-3PW-01-1DWDUP	966	0.1049	-	7470A	PPB	23 Jan 2019 14:10:56	U	SMPL
K1218-01MS - 1	4-3PW-01-1DWMS	18581	4.1601	-	7470A	PPB	23 Jan 2019 14:13:04	U	SMPL
K1218-01MSD - 1	4-3PW-01-1DWMSD	16986	3.7929	-	7470A	PPB	23 Jan 2019 14:15:12	U	SMPL
PB116579TB - 1	PB116555TB	614	0.0239	-	7470A	PPB	23 Jan 2019 14:17:20	U	SMPL
CCV08 - 1	CCV08	22520	5.0669	-	7470A	PPB	23 Jan 2019 14:19:28	U	SMPL
CCB08 - 1	CCB08	844	0.0768	-	7470A	PPB	23 Jan 2019 14:21:38	U	SMPL
K1210-02LX5 - 1		755	0.0564	-	7470A	PPB	23 Jan 2019 14:23:45	U	SMPL
K1210-02A - 1		15965	3.5578	-	7470A	PPB	23 Jan 2019 14:25:53	U	SMPL
K1218-01LX5 - 1		792	0.0649	-	7470A	PPB	23 Jan 2019 14:28:00	U	SMPL
K1218-01A - 1		17049	3.8074	-	7470A	PPB	23 Jan 2019 14:30:10	U	SMPL
CCV09 - 1	CCV09	20879	4.6891	-	7470A	PPB	23 Jan 2019 14:36:43	U	SMPL
CCB09 - 1	CCB09	566	0.0128	-	7470A	PPB	23 Jan 2019 14:38:50	U	SMPL