



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100452

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	01/23/2019	15:00
2	CAL2	1	Water	NA	NA	21.3	7.01	01/23/2019	15:01
3	CAL3	1	Water	NA	NA	21.3	10.03	01/23/2019	15:02
4	ICV	1	Water	NA	NA	21.3	6.99	01/23/2019	15:03
5	CCV1	1	Water	NA	NA	21.2	2.02	01/23/2019	15:05
6	K1223-01	1	Solid	20.02	20	20.7	8.56	01/23/2019	15:06
7	K1223-02	1	Solid	20.03	20	20.5	8.46	01/23/2019	15:07
8	K1223-03	1	Solid	20.02	20	20.4	8.46	01/23/2019	15:08
9	K1223-04	1	Solid	20.03	20	20.7	8.57	01/23/2019	15:10
10	K1223-05	1	Solid	20.04	20	20.4	8.46	01/23/2019	15:11
11	K1223-06	1	Solid	20.02	20	20.6	8.66	01/23/2019	15:13
12	K1223-07	1	Solid	20.03	20	20.2	8.36	01/23/2019	15:15
13	K1223-08	1	Solid	20.05	20	20.5	8.55	01/23/2019	15:16
14	K1223-09	1	Solid	20.02	20	20.2	8.52	01/23/2019	15:18
15	K1223-10	1	Solid	20.03	20	20.5	8.68	01/23/2019	15:20
16	CCV2	1	Water	NA	NA	21.1	12.01	01/23/2019	15:21
17	K1223-10DUP	1	Solid	20.05	20	20.4	8.69	01/23/2019	15:23
18	CCV3	1	Water	NA	NA	21.2	2.02	01/23/2019	15:24

203100452

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1223

WorkList ID : 121569

Date : 1/23/2019 2:49:55 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/23/2019	Solid	K1223-01 - 1	pH	Cool 4 deg C	ENTA05	L12	TS-M	01/23/2019	9045D
01/23/2019	Solid	K1223-02 - 1	pH	Cool 4 deg C	ENTA05	L12	TS-N	01/23/2019	9045D
01/23/2019	Solid	K1223-03 - 1	pH	Cool 4 deg C	ENTA05	L12	TS-O	01/23/2019	9045D
01/23/2019	Solid	K1223-04 - 1	pH	Cool 4 deg C	ENTA05	L12	TS-P	01/23/2019	9045D
01/23/2019	Solid	K1223-05 - 1	pH	Cool 4 deg C	ENTA05	L12	TS-Q	01/23/2019	9045D
01/23/2019	Solid	K1223-06 - 1	pH	Cool 4 deg C	ENTA05	L12	TS-R	01/23/2019	9045D
01/23/2019	Solid	K1223-07 - 1	pH	Cool 4 deg C	ENTA05	L12	TS-S	01/23/2019	9045D
01/23/2019	Solid	K1223-08 - 1	pH	Cool 4 deg C	ENTA05	L12	TS-T	01/23/2019	9045D
01/23/2019	Solid	K1223-09 - 1	pH	Cool 4 deg C	ENTA05	L12	TS-U	01/23/2019	9045D
01/23/2019	Solid	K1223-10 - 1	pH	Cool 4 deg C	ENTA05	L12	TS-V	01/23/2019	9045D

Date/Time 01-23-19 2:50 PM

Received by: JD

Relinquished by: [Signature]

Date/Time 01-23-19 4:25 PM

Received by: [Signature]

Relinquished by: JD