

LB100460


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

47521



A= 0.0000e+000

B= 2.0534e-004

C= 6.2039e-002

Rho= 0.9995528

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.062	-0.062	-606	0.000	-606				
0.2	0.200	0.166	-0.034	506	0.0 %	506				
2.5	2.500	2.516	0.016	11950	0.0 %	11950				
5.0	5.000	5.104	0.104	24555	0.0 %	24555				
7.5	7.500	7.657	0.157	36985	0.0 %	36985				
10.0	10.000	9.820	-0.180	47521	0.0 %	47521				

LB100460 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Jan 2019 10:56:14

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-606	-	0.0000	7471B	PPB	24 Jan 2019 15:35:17	S	Std
0.2 - 1	50.2	506	-	0.2000	7471B	PPB	24 Jan 2019 15:37:36	S	Std
2.5 - 1	52.5	11950	-	2.5000	7471B	PPB	24 Jan 2019 15:39:54	S	Std
5.0 - 1	55.0	24555	-	5.0000	7471B	PPB	24 Jan 2019 15:42:12	S	Std
7.5 - 1	57.5	36985	-	7.5000	7471B	PPB	24 Jan 2019 15:44:30	S	Std
10.0 - 1	510.0	47521	-	10.0000	7471B	PPB	24 Jan 2019 15:46:47	S	Std
ICV78 - 1	ICV78	18950	3.9532	-	7471B	PPB	24 Jan 2019 15:58:20	U	SMPL
ICB78 - 1	ICB78	173	0.0976	-	7471B	PPB	24 Jan 2019 16:00:36	U	SMPL
CCV10 - 1	CCV10	24026	4.9955	-	7471B	PPB	24 Jan 2019 16:23:09	U	SMPL
CCB10 - 1	CCB10	6	0.0633	-	7471B	PPB	24 Jan 2019 16:25:26	U	SMPL
CRA - 1	CRA	944	0.2559	-	7471B	PPB	24 Jan 2019 16:54:46	U	SMPL
HighStd - 1	HighStd	49825	10.2931	-	7471B	PPB	24 Jan 2019 16:57:02	U	SMPL
ChkStd - 1	ChkStd	34627	7.1723	-	7471B	PPB	24 Jan 2019 16:59:18	U	SMPL
PB116598BL - 1	PBS	38	0.0698	-	7471B	PPB	24 Jan 2019 17:01:34	U	SMPL
PB116598BS - 1	LCSS	18578	3.8768	-	7471B	PPB	24 Jan 2019 17:03:51	U	SMPL
K1003-01 - 1	BU-02-012319	2202	0.5142	-	7471B	PPB	24 Jan 2019 17:06:08	U	SMPL
K1003-01DUP - 1	BU-02-012319DUP	2338	0.5421	-	7471B	PPB	24 Jan 2019 17:08:25	U	SMPL
K1003-01MS - 1	BU-02-012319MS	19997	4.1682	-	7471B	PPB	24 Jan 2019 17:15:27	U	SMPL
K1003-01MSD - 1	BU-02-012319MSD	18930	3.9491	-	7471B	PPB	24 Jan 2019 17:17:44	U	SMPL
K1016-03 - 1	TR-05-RO-1-012319	5931	1.2799	-	7471B	PPB	24 Jan 2019 17:20:00	U	SMPL
CCV11 - 1	CCV11	24484	5.0896	-	7471B	PPB	24 Jan 2019 17:22:16	U	SMPL
CCB11 - 1	CCB11	139	0.0906	-	7471B	PPB	24 Jan 2019 17:24:32	U	SMPL
K1016-05 - 1	TR-05-RO-2-012319	2008	0.4744	-	7471B	PPB	24 Jan 2019 17:26:48	U	SMPL
K1016-07 - 1	TR-05-RO-3-012319	3468	0.7742	-	7471B	PPB	24 Jan 2019 17:29:05	U	SMPL
K1016-09 - 1	TR-05-RO-4-012319	1785	0.4286	-	7471B	PPB	24 Jan 2019 17:31:22	U	SMPL
K1223-01 - 1	TS-M	6548	1.4066	-	7471B	PPB	24 Jan 2019 17:33:39	U	SMPL
K1223-02 - 1	TS-N	7326	1.5664	-	7471B	PPB	24 Jan 2019 17:35:55	U	SMPL
K1223-03 - 1	TS-O	7275	1.5559	-	7471B	PPB	24 Jan 2019 17:38:13	U	SMPL
K1223-04 - 1	TS-P	8002	1.7052	-	7471B	PPB	24 Jan 2019 17:40:30	U	SMPL
K1223-05 - 1	TS-Q	7512	1.6045	-	7471B	PPB	24 Jan 2019 17:42:49	U	SMPL
K1223-06 - 1	TS-R	7180	1.5364	-	7471B	PPB	24 Jan 2019 17:45:07	U	SMPL
K1223-07 - 1	TS-S	8342	1.7750	-	7471B	PPB	24 Jan 2019 17:47:23	U	SMPL
CCV12 - 1	CCV12	25550	5.3085	-	7471B	PPB	24 Jan 2019 17:49:40	U	SMPL
CCB12 - 1	CCB12	45	0.0713	-	7471B	PPB	24 Jan 2019 17:51:56	U	SMPL
K1223-08 - 1	TS-T	11921	2.5099	-	7471B	PPB	24 Jan 2019 17:54:13	U	SMPL
K1223-09 - 1	TS-U	7568	1.6160	-	7471B	PPB	24 Jan 2019 17:56:30	U	SMPL
K1223-10 - 1	TS-V	5525	1.1965	-	7471B	PPB	24 Jan 2019 17:58:47	U	SMPL
K1003-01LX5 - 1		382	0.1405	-	7471B	PPB	24 Jan 2019 18:01:03	U	SMPL
K1003-01A - 1		22798	4.7434	-	7471B	PPB	24 Jan 2019 18:11:51	U	SMPL
CCV13 - 1	CCV13	23734	4.9356	-	7471B	PPB	24 Jan 2019 18:18:05	U	SMPL
CCB13 - 1	CCB13	642	0.1939	-	7471B	PPB	24 Jan 2019 18:20:22	U	SMPL