



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100466

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	01/24/2019	15:45
2	CAL2	1	Water	NA	NA	21.2	7.00	01/24/2019	15:46
3	CAL3	1	Water	NA	NA	21.2	10.03	01/24/2019	15:47
4	ICV	1	Water	NA	NA	21.1	6.99	01/24/2019	15:48
5	CCV1	1	Water	NA	NA	21.1	2.02	01/24/2019	15:49
6	K1236-01	1	Solid	20.03	20	20.8	6.60	01/24/2019	15:50
7	K1236-01DUP	1	Solid	20.04	20	20.7	6.61	01/24/2019	15:51
8	K1236-02	1	Solid	20.05	20	20.5	8.20	01/24/2019	16:00
9	K1236-03	1	Solid	20.03	20	20.7	8.15	01/24/2019	16:02
10	CCV2	1	Water	NA	NA	21.1	12.01	01/24/2019	16:03

LB 100461

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1236

WorkList ID : 121602

Date : 1/24/2019 3:36:42 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/24/2019	Solid	K1236-01 ✓	pH	Cool 4 deg C	ENVI77	L22	ROLLOFF-263	01/24/2019	9045D
01/24/2019	Solid	K1236-02 ✓	pH	Cool 4 deg C	ENVI77	L22	ROLLOFF-274	01/24/2019	9045D
01/24/2019	Solid	K1236-03 ✓	pH	Cool 4 deg C	ENVI77	L22	ROLLOFF-284	01/24/2019	9045D

Date/Time 01-24-19 3:40 PM

Received by: Jn

Relinquished by: Cg

Date/Time 01-24-19 4:30 PM

Received by: Cg

Relinquished by: Jn