

LB100472


245.1

INSTRUMENT ID : CV1

Linear

 μ Abs. :

42392



A= 0.0000e+000

B= 2.3737e-004

C= 3.3804e-003

Rho= 0.9999333

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.005	0.005	5	0.000	5	0			
0.05	0.050	0.058	0.008	229	0.0 %	229				
0.2	0.200	0.211	0.011	874	0.0 %	874				
2.5	2.500	2.500	0.000	10519	0.0 %	10519				
5.0	5.000	5.002	0.002	21059	0.0 %	21059				
7.5	7.500	7.409	-0.091	31198	0.0 %	31198				
10.0	10.000	10.066	0.066	42392	0.0 %	42392				

LB100472 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 25 Jan 2019 09:57:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	5	-	0.0000	245.1	PPB	25 Jan 2019 09:59:16	S	Std
0.05 - 1	50.05	229	-	0.0500	245.1	PPB	25 Jan 2019 10:05:37	S	Std
0.2 - 1	50.2	874	-	0.2000	245.1	PPB	25 Jan 2019 10:07:55	S	Std
2.5 - 1	52.5	10519	-	2.5000	245.1	PPB	25 Jan 2019 10:10:13	S	Std
5.0 - 1	55.0	21059	-	5.0000	245.1	PPB	25 Jan 2019 10:12:31	S	Std
7.5 - 1	57.5	31198	-	7.5000	245.1	PPB	25 Jan 2019 10:14:47	S	Std
10.0 - 1	510.0	42392	-	10.0000	245.1	PPB	25 Jan 2019 10:23:26	S	Std
ICV79 - 1	ICV79	16270	3.8653	-245.1	PPB	PPB	25 Jan 2019 10:27:28	U	SMPL
ICB79 - 1	ICB79	-17	-0.0007	-245.1	PPB	PPB	25 Jan 2019 10:29:44	U	SMPL
CCV14 - 1	CCV14	21098	5.0113	-245.1	PPB	PPB	25 Jan 2019 10:32:00	U	SMPL
CCB14 - 1	CCB14	-52	-0.0090	-245.1	PPB	PPB	25 Jan 2019 10:34:15	U	SMPL
CRA - 1	CRA	236	0.0594	-245.1	PPB	PPB	25 Jan 2019 10:36:31	U	SMPL
HighStd - 1	HighStd	41907	9.9507	-245.1	PPB	PPB	25 Jan 2019 10:38:47	U	SMPL
ChkStd - 1	ChkStd	29275	6.9523	-245.1	PPB	PPB	25 Jan 2019 10:41:04	U	SMPL
PB116601BL - 1	PBW	-53	-0.0092	-245.1	PPB	PPB	25 Jan 2019 10:43:21	U	SMPL
PB116601BS - 1	LCSW	16310	3.8748	-245.1	PPB	PPB	25 Jan 2019 10:45:38	U	SMPL
K1161-05 - 1	PT-HG-WS	8553	2.0336	-245.1	PPB	PPB	25 Jan 2019 10:47:55	U	SMPL
K1206-01 - 1	011819-DBRI-E-D-M	101	0.0274	-245.1	PPB	PPB	25 Jan 2019 10:50:12	U	SMPL
K1206-02 - 1	011819-DBRI-E-U-M	110	0.0295	-245.1	PPB	PPB	25 Jan 2019 11:03:09	U	SMPL
K1206-02DUP - 1	011819-DBRI-E-U-MDUP	101	0.0274	-245.1	PPB	PPB	25 Jan 2019 11:05:26	U	SMPL
K1206-02MS - 1	011819-DBRI-E-U-MMS	17224	4.0918	-245.1	PPB	PPB	25 Jan 2019 11:07:42	U	SMPL
CCV15 - 1	CCV15	20619	4.8976	-245.1	PPB	PPB	25 Jan 2019 11:09:58	U	SMPL
CCB15 - 1	CCB15	-33	-0.0045	-245.1	PPB	PPB	25 Jan 2019 11:12:14	U	SMPL
K1206-02MSD - 1	011819-DBRI-E-U-MMSD	17788	4.2257	-245.1	PPB	PPB	25 Jan 2019 11:14:31	U	SMPL
K1206-02LX5 - 1		42	0.0133	-245.1	PPB	PPB	25 Jan 2019 11:16:47	U	SMPL
K1206-02A - 1		19761	4.6940	-245.1	PPB	PPB	25 Jan 2019 11:19:02	U	SMPL
CCV16 - 1	CCV16	20547	4.8805	-245.1	PPB	PPB	25 Jan 2019 11:33:09	U	SMPL
CCB16 - 1	CCB16	-59	-0.0106	-245.1	PPB	PPB	25 Jan 2019 11:35:27	U	SMPL