



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100473

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/25/2019	10:35
2	CAL2	1	Water	NA	NA	20.1	7.00	01/25/2019	10:36
3	CAL3	1	Water	NA	NA	20.1	10.02	01/25/2019	10:37
4	ICV	1	Water	NA	NA	20.1	7.00	01/25/2019	10:38
5	CCV1	1	Water	NA	NA	20.2	2.01	01/25/2019	10:40
6	K1247-01	1	Water	NA	NA	10.2	4.52	01/25/2019	10:41
7	K1247-01DUP	1	Water	NA	NA	10.3	4.53	01/25/2019	10:42
8	K1247-02	1	Water	NA	NA	10.1	4.21	01/25/2019	10:43
9	K1247-03	1	Water	NA	NA	10.3	5.07	01/25/2019	10:44
10	CCV2	1	Water	NA	NA	20.1	12.03	01/25/2019	10:45

LB100473

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph k1247

WorkList ID : 121626

Date : 1/25/2019 10:31:46 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/24/2019	Water	K1247-01 -VA	pH	Cool 4 deg C	STER09	L23	CELL-A-POOL-WATER(DE-IC)	01/24/2019	SM 4500-H B
01/24/2019	Water	K1247-02 -VA	pH	Cool 4 deg C	STER09	L23	CELL-B-C-POOL-WATER(DE-IC)	01/24/2019	SM 4500-H B
01/24/2019	Water	K1247-03 -VA	pH	Cool 4 deg C	STER09	L23	CELL-D-POOL-WATER(DE-IC)	01/24/2019	SM 4500-H B

Date/Time 01-25-19 10:32 AM

Received by: Jn

Relinquished by: CO

Date/Time 01-25-19 11:15 AM

Received by: CO

Relinquished by: JP