



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB100475

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP71193
Turbidity std, 40NTU	WP71187
turbidity std 5 NTU	WP71190
turbidity std 20 NTU	WP71188
Turbidity std 1 NTU	WP71191
turbidity std 0 NTU	WP71192
turbidity std 10 NTU	WP71189
10 NTU Standard 500 ml	W2297

Intercept: 0.0913

Slope: 0.9314

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.247	01/25/2019	10:07
2	CAL2	1	1	1.220	01/25/2019	10:09
3	CAL3	5	1	4.550	01/25/2019	10:11
4	CAL4	10	1	9.320	01/25/2019	10:13
5	CAL5	20	1	18.500	01/25/2019	10:14
6	CAL6	40	1	37.500	01/25/2019	10:17



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB100475

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.100	10.746	01/25/2019	10:18
2	ICB		1	0.235	0.154	01/25/2019	10:20
3	CCV1	10	1	9.400	9.994	01/25/2019	10:22
4	CCB1		1	0.243	0.163	01/25/2019	10:24
5	LB100475BL		1	0.245	0.165	01/25/2019	10:26
6	K1161-01		1	1.210	1.201	01/25/2019	10:29
7	K1161-01DUP		1	1.230	1.223	01/25/2019	10:32
8	CCV2	10	1	9.380	9.973	01/25/2019	10:34
9	CCB2		1	0.240	0.160	01/25/2019	10:36

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tur011719

WorkList ID : 121414

Date : 1/17/2019 1:02:23 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/16/2019	Water	K1161-01	Turbidity	Cool 4 deg C	CHEM02	QA Office	PT-TURB-WS <u>B100475</u>	01/14/2019	180.1
01/16/2019	Water	K1161-02	Turbidity	Cool 4 deg C	CHEM02	QA Office	PT-TURB-WS <u>B100476</u>	01/14/2019	SM2130 B

Date/Time 1/17/19 1:15 PM
 Received by: AS
 Relinquished by: MD

Date/Time 1/25/19 12:30 PM
 Received by: AS
 Relinquished by: AS