

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB100484

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP70044
calibration std. phosphate 1 ppm	WP71201
calibration std. phosphate 0.5 ppm	WP71200
calibration std. phosphate 0.3 ppm	WP71199
calibration std. 0 ppm	WP71196
phosphate ICV std.	WP71203
phosphate CCV std.	WP71202
calibration std. phosphate 0.1 ppm	WP71198
calibration std. phosphate 0.05 ppm	WP71197
Combined reagent	WP71206
11 N sulfuric acid	WP66905
Phenolphthalein indicator	WP66907
Sodium hydroxide, 1N	WP66924

Intercept: -0.0008Slope: 0.6506Regression: 0.99981

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	01/25/2019	15:02
2	CAL2	0.05	1	50	50	0.033	01/25/2019	15:03
3	CAL3	0.10	1	50	50	0.065	01/25/2019	15:03
4	CAL4	0.30	1	50	50	0.186	01/25/2019	15:04
5	CAL5	0.50	1	50	50	0.331	01/25/2019	15:04
6	CAL6	1.00	1	50	50	0.649	01/25/2019	15:05



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB100484

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.324	0.499	01/25/2019	15:05
2	ICB		1	50	50	0.000	0.001	01/25/2019	15:06
3	CCV1	0.50	1	50	50	0.328	0.505	01/25/2019	15:07
4	CCB1		1	50	50	0.000	0.001	01/25/2019	15:07
5	PB116636BL		1	50	50	0.000	0.001	01/25/2019	15:08
6	PB116636BS	0.50	1	50	50	0.303	0.467	01/25/2019	15:08
7	K1249-01		1	50	50	0.621	0.956	01/25/2019	15:09
8	K1249-01DUP		1	50	50	0.557	0.857	01/25/2019	15:09
9	K1249-01MS	0.50	1	50	50	0.828	1.274	01/25/2019	15:10
10	K1249-01MSD	0.50	1	50	50	0.830	1.277	01/25/2019	15:11
11	K1249-01MS	0.50	2	50	50	0.436	0.671	01/25/2019	15:12
12	K1249-01MSD	0.50	2	50	50	0.448	0.690	01/25/2019	15:12
13	CCV2	0.50	1	50	50	0.326	0.502	01/25/2019	15:13
14	CCB2		1	50	50	0.000	0.001	01/25/2019	15:13