



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100490

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	01/25/2019	16:25
2	CAL2	1	Water	NA	NA	21.2	7.01	01/25/2019	16:26
3	CAL3	1	Water	NA	NA	21.2	10.03	01/25/2019	16:27
4	ICV	1	Water	NA	NA	21.1	6.99	01/25/2019	16:28
5	CCV1	1	Water	NA	NA	21.2	2.01	01/25/2019	16:30
6	K1250-05	1	Water	NA	NA	16.0	7.47	01/25/2019	16:31
7	K1250-05DUP	1	Water	NA	NA	16.1	7.48	01/25/2019	16:32
8	CCV2	1	Water	NA	NA	21.2	12.02	01/25/2019	16:33

2B100490

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrsoivity k1257

WorkList ID : 121640

Date : 1/25/2019 4:14:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/25/2019	Water	AK1250-05	Corrosivity	Cool 4 deg C	PSEG01	K31	DRUM-1-3	01/25/2019	9040C

Date/Time 01-25-19 4:20 PM

Received by: JD

Relinquished by: CD

Date/Time 01-25-19 5:35 PM

Received by: SJ

Relinquished by: JP