



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB100491

Analyst By : jignesh

Supervisor Review By : apatel

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	01/25/2019	16:25
2	CAL2	1	Water	NA	NA	21.2	7.01	01/25/2019	16:26
3	CAL3	1	Water	NA	NA	21.2	10.03	01/25/2019	16:27
4	ICV	1	Water	NA	NA	21.1	6.99	01/25/2019	16:28
5	CCV1	1	Water	NA	NA	21.2	2.01	01/25/2019	16:30
6	K1250-04	1	Water	NA	NA	16.0	7.47	01/25/2019	16:31
7	K1250-04DUP	1	Water	NA	NA	16.1	7.48	01/25/2019	16:32
8	K1256-01	1	Water	NA	NA	13.8	6.59	01/25/2019	16:34
9	K1256-02	1	Water	NA	NA	13.9	5.53	01/25/2019	16:35
10	K1256-03	1	Water	NA	NA	14.7	6.74	01/25/2019	16:37
11	K1256-04	1	Water	NA	NA	14.4	5.65	01/25/2019	16:39
12	K1256-05	1	Water	NA	NA	14.9	6.30	01/25/2019	16:40
13	K1256-06	1	Water	NA	NA	15.3	6.72	01/25/2019	16:41
14	CCV2	1	Water	NA	NA	21.2	12.02	01/25/2019	16:42

LB100491

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph k1257

WorkList ID : 121641

Date : 1/25/2019 4:15:15 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/25/2019	Water	K1250-04 - 1A	pH	Cool 4 deg C	PSEG01	K31	DRUM-1-3	01/25/2019	9040C
01/25/2019	Water	K1256-01 - C	pH	Cool 4 deg C	PSEG01	K31	30093	01/25/2019	9040C
01/25/2019	Water	K1256-02 - C	pH	Cool 4 deg C	PSEG01	K31	30175	01/25/2019	9040C
01/25/2019	Water	K1256-03 - C	pH	Cool 4 deg C	PSEG01	K31	30176	01/25/2019	9040C
01/25/2019	Water	K1256-04 - C	pH	Cool 4 deg C	PSEG01	K31	50138	01/25/2019	9040C
01/25/2019	Water	K1256-05 - C	pH	Cool 4 deg C	PSEG01	K31	50139	01/25/2019	9040C
01/25/2019	Water	K1256-06 - C	pH	Cool 4 deg C	PSEG01	K31	50140	01/25/2019	9040C

Date/Time 01-25-19 6:12:00 PM

Received by: Jy

Relinquished by: CD

Date/Time

Received by: LG

Relinquished by: JP

01-25-19 5:35 PM