

LB100503


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

60528



A= 0.0000e+000

B= 1.6614e-004

C= -2.6974e-003

Rho= 0.9998961

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	24	0.000	24				
0.2	0.200	0.200	-0.000	1219	0.0 %	1219				
2.5	2.500	2.504	0.004	15087	0.0 %	15087				
5.0	5.000	5.049	0.049	30407	0.0 %	30407				
7.5	7.500	7.392	-0.108	44508	0.0 %	44508				
10.0	10.000	10.054	0.054	60528	0.0 %	60528				

LB100503 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 Jan 2019 10:16:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	24	-	0.00007471B	PPB		28 Jan 2019 10:20:01	S	Std
0.2 - 1	50.2	1219	-	0.20007471B	PPB		28 Jan 2019 10:22:19	S	Std
2.5 - 1	52.5	15087	-	2.50007471B	PPB		28 Jan 2019 10:24:38	S	Std
5.0 - 1	55.0	30407	-	5.00007471B	PPB		28 Jan 2019 10:26:55	S	Std
7.5 - 1	57.5	44508	-	7.50007471B	PPB		28 Jan 2019 10:29:12	S	Std
10.0 - 1	510.0	60528	-	10.00007471B	PPB		28 Jan 2019 10:31:29	S	Std
ICV81 - 1	ICV81	24582	4.0815	-7471B	PPB		28 Jan 2019 10:35:23	U	SMPL
ICB81 - 1	ICB81	-50	-0.0110	-7471B	PPB		28 Jan 2019 10:37:38	U	SMPL
CCV20 - 1	CCV20	30882	5.1282	-7471B	PPB		28 Jan 2019 10:39:54	U	SMPL
CCB20 - 1	CCB20	-71	-0.0145	-7471B	PPB		28 Jan 2019 10:42:10	U	SMPL
CRA - 1	CRA	1191	0.1952	-7471B	PPB		28 Jan 2019 10:44:27	U	SMPL
HighStd - 1	HighStd	60825	10.1030	-7471B	PPB		28 Jan 2019 10:46:43	U	SMPL
ChkStd - 1	ChkStd	43612	7.2432	-7471B	PPB		28 Jan 2019 10:52:47	U	SMPL
PB116637BL - 1	PBS	-105	-0.0201	-7471B	PPB		28 Jan 2019 10:55:03	U	SMPL
PB116637BS - 1	LCSS	21504	3.5701	-7471B	PPB		28 Jan 2019 10:57:21	U	SMPL
K1015-13 - 1	SU-04-012419-A	1474	0.2422	-7471B	PPB		28 Jan 2019 10:59:39	U	SMPL
K1015-15 - 1	SU-04-012419-B	2142	0.3532	-7471B	PPB		28 Jan 2019 11:01:57	U	SMPL
K1228-01 - 1	COMP-1	14142	2.3469	-7471B	PPB		28 Jan 2019 11:04:14	U	SMPL
K1228-01DUP - 1	COMP-1DUP	14140	2.3466	-7471B	PPB		28 Jan 2019 11:06:32	U	SMPL
K1228-01MS - 1	COMP-1MS	24174	4.0137	-7471B	PPB		28 Jan 2019 11:08:49	U	SMPL
CCV21 - 1	CCV21	31509	5.2323	-7471B	PPB		28 Jan 2019 11:11:05	U	SMPL
CCB21 - 1	CCB21	-94	-0.0183	-7471B	PPB		28 Jan 2019 11:13:22	U	SMPL
K1228-01MSD - 1	COMP-1MSD	24268	4.0293	-7471B	PPB		28 Jan 2019 11:15:41	U	SMPL
K1231-01 - 1	RBR-200029	5401	0.8946	-7471B	PPB		28 Jan 2019 11:17:58	U	SMPL
K1233-01 - 1	SANDHILL-RO-1-8	406	0.0648	-7471B	PPB		28 Jan 2019 11:20:15	U	SMPL
K1234-03 - 1	EWP	15037	2.4956	-7471B	PPB		28 Jan 2019 11:22:31	U	SMPL
K1237-01 - 1	BCB-COMPOSITE	591	0.0955	-7471B	PPB		28 Jan 2019 11:24:47	U	SMPL
K1238-01 - 1	S1	1282	0.2103	-7471B	PPB		28 Jan 2019 11:27:05	U	SMPL
K1238-11 - 1	S2	1226	0.2010	-7471B	PPB		28 Jan 2019 11:29:22	U	SMPL
K1238-21 - 1	S3	777	0.1264	-7471B	PPB		28 Jan 2019 11:31:39	U	SMPL
K1238-31 - 1	S4	1310	0.2150	-7471B	PPB		28 Jan 2019 11:33:58	U	SMPL
K1238-41 - 1	S5	839	0.1367	-7471B	PPB		28 Jan 2019 11:36:14	U	SMPL
CCV22 - 1	CCV22	30649	5.0895	-7471B	PPB		28 Jan 2019 11:38:30	U	SMPL
CCB22 - 1	CCB22	-82	-0.0163	-7471B	PPB		28 Jan 2019 11:40:47	U	SMPL
K1240-01 - 1	390109125	82228	13.6590	-7471B	PPB		28 Jan 2019 11:43:05	U	SMPL
K1241-01 - 1	390109124	5612	0.9297	-7471B	PPB		28 Jan 2019 11:45:21	U	SMPL
K1228-01LX5 - 1		2737	0.4520	-7471B	PPB		28 Jan 2019 11:47:40	U	SMPL
K1228-01A - 1		38369	6.3721	-7471B	PPB		28 Jan 2019 11:49:57	U	SMPL
K1240-01DLX2 - 1		42506	7.0594	-7471B	PPB		28 Jan 2019 11:52:14	U	SMPL
CCV23 - 1	CCV23	31484	5.2282	-7471B	PPB		28 Jan 2019 11:54:32	U	SMPL
CCB23 - 1	CCB23	-93	-0.0181	-7471B	PPB		28 Jan 2019 11:56:50	U	SMPL